

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/21/2025	BNA_F	<u>BF143540.D</u>
Compound Name	Response	Retention Time
DDT	210961	13.61
DDD		13.31
DDE	530	12.974
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
530	211491	0.25