

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\  
 Data File : BF108382.D  
 Acq On : 22 Aug 2018 14:49  
 Operator : JU/SJ  
 Sample : PB112238BL  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB112238BL

Quant Time: Aug 22 16:34:28 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 22 11:01:45 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.00  | 152  | 166162   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.28  | 136  | 631485   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 10.04 | 164  | 332799   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.54 | 188  | 691843   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 14.19 | 240  | 436779   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.75 | 264  | 312833   | 20.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.64  | 112  | 1177477  | 128.29 | ng    | 0.01     |
| 7) Phenol-d6                | 6.62  | 99   | 1568824  | 128.63 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.57  | 82   | 1066712  | 94.26  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.84 | 330  | 529141   | 159.40 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.36  | 172  | 2062470  | 99.50  | ng    | 0.00     |
| 78) Terphenyl-d14           | 13.13 | 244  | 2227729  | 129.68 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\  
Data File : BF108382.D  
Acq On : 22 Aug 2018 14:49  
Operator : JU/SJ  
Sample : PB112238BL  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB112238BL

Quant Time: Aug 22 16:34:28 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 22 11:01:45 2018  
Response via : Initial Calibration

