

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\
 Data File : BF108399.D
 Acq On : 22 Aug 2018 22:33
 Operator : JU/SJ
 Sample : J4573-11 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T27-COMP

Manual Integrations
 APPROVED

Sohil
 8/23/2018 3:43:29 PM

Quant Time: Aug 23 12:31:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 11:01:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.00	152	154256	20.00	ng	0.00
21) Naphthalene-d8	8.28	136	583325	20.00	ng	0.00
38) Acenaphthene-d10	10.05	164	268839	20.00	ng	0.00
63) Phenanthrene-d10	11.54	188	439454	20.00	ng	0.00
75) Chrysene-d12	14.19	240	302337	20.00	ng	0.00
86) Perylene-d12	15.75	264	260687	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	129572	15.21	ng	-0.01
7) Phenol-d6	6.61	99	176921	15.63	ng	-0.02
23) Nitrobenzene-d5	7.56	82	107201	10.25	ng	-0.01
41) 2,4,6-Tribromophenol	10.84	330	28585	10.66	ng	-0.01
44) 2-Fluorobiphenyl	9.35	172	186831	11.16	ng	-0.01
78) Terphenyl-d14	13.12	244	133694	11.24	ng	0.00
Target Compounds						
74) Fluoranthene	12.76	202	45796	2.024	ng	94
77) Pyrene	12.99	202	85865	4.486	ng	97
85) Indeno(1,2,3-cd)pyrene	17.35	276	33501	2.431	ng	# 90
87) Benzo(b)fluoranthene	15.28	252	55445m	3.559	ng	
89) Benzo(a)pyrene	15.68	252	65554m	4.700	ng	
91) Benzo(g,h,i)perylene	17.85	276	48442	4.263	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\
Data File : BF108399.D
Acq On : 22 Aug 2018 22:33
Operator : JU/SJ
Sample : J4573-11 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
T27-COMP

Quant Time: Aug 23 12:31:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 11:01:45 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
8/23/2018 3:43:29 PM

