

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\  
 Data File : BF108411.D  
 Acq On : 23 Aug 2018 5:47  
 Operator : JU/SJ  
 Sample : J4525-12  
 Misc :  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 081518-SIDI-F-U-B

Quant Time: Aug 23 08:30:15 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 22 11:01:45 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 7.00  | 152  | 149996   | 20.00  | ng    | 0.00           |
| 21) Naphthalene-d8          | 8.28  | 136  | 538686   | 20.00  | ng    | 0.00           |
| 38) Acenaphthene-d10        | 10.05 | 164  | 252429   | 20.00  | ng    | 0.00           |
| 63) Phenanthrene-d10        | 11.54 | 188  | 420165   | 20.00  | ng    | 0.00           |
| 75) Chrysene-d12            | 14.19 | 240  | 289202   | 20.00  | ng    | -0.01          |
| 86) Perylene-d12            | 15.75 | 264  | 239317   | 20.00  | ng    | -0.01          |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.62  | 112  | 335795   | 40.53  | ng    | 0.00           |
| 7) Phenol-d6                | 6.61  | 99   | 270877   | 24.60  | ng    | -0.02          |
| 23) Nitrobenzene-d5         | 7.57  | 82   | 972482   | 100.74 | ng    | 0.00           |
| 41) 2,4,6-Tribromophenol    | 10.84 | 330  | 188128   | 74.72  | ng    | -0.01          |
| 44) 2-Fluorobiphenyl        | 9.36  | 172  | 1725545  | 109.75 | ng    | 0.00           |
| 78) Terphenyl-d14           | 13.13 | 244  | 1249113  | 109.82 | ng    | 0.00           |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 9.75  | 163  | 72604    | 4.129  | ng    | Qvalue<br># 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082218\  
Data File : BF108411.D  
Acq On : 23 Aug 2018 5:47  
Operator : JU/SJ  
Sample : J4525-12  
Misc :  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
081518-SIDI-F-U-B

Quant Time: Aug 23 08:30:15 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 22 11:01:45 2018  
Response via : Initial Calibration

