

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082316\
 Data File : BF089822.D
 Acq On : 23 Aug 2016 9:39
 Operator : UM/SJ
 Sample : H4547-01
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 D-1(2-2.5)

Quant Time: Aug 25 13:18:53 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 19 14:02:57 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	473434	20.00	ng	-0.02
21) Naphthalene-d8	7.95	136	1956618	20.00	ng	-0.02
38) Acenaphthene-d10	9.70	164	931819	20.00	ng	-0.01
63) Phenanthrene-d10	11.18	188	1700753	20.00	ng	-0.01
75) Chrysene-d12	13.91	240	1583017	20.00	ng	0.03
86) Perylene-d12	15.49	264	1475831	20.00	ng	0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	3165018	114.64	ng	-0.01
7) Phenol-d6	6.31	99	4088386	120.64	ng	-0.01
23) Nitrobenzene-d5	7.23	82	2302089	73.09	ng	-0.02
41) 2,4,6-Tribromophenol	10.49	330	983237	111.03	ng	-0.01
44) 2-Fluorobiphenyl	9.04	172	4230909	79.80	ng	-0.01
78) Terphenyl-d14	12.79	244	3778061	53.99	ng	0.00
Target Compounds						
10) Phenol	6.32	94	120308	3.03	ng	# 75
31) Naphthalene	7.98	128	2862940	31.31	ng	99
37) 2-Methylnaphthalene	8.66	142	927784	15.68	ng	96
49) Dimethylphthalate	9.44	163	1407924	21.46	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082316\
Data File : BF089822.D
Acq On : 23 Aug 2016 9:39
Operator : UM/SJ
Sample : H4547-01
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
D-1(2-2.5)

Quant Time: Aug 25 13:18:53 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 19 14:02:57 2016
Response via : Initial Calibration

