

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF082316\  
 Data File : BF089861.D  
 Acq On : 24 Aug 2016 5:08  
 Operator : UM/SJ  
 Sample : BL-CHIPS-4  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 BL-CHIPS-4

Quant Time: Aug 25 14:32:29 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF081916.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Aug 19 14:02:57 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.66  | 152  | 439943   | 20.00  | ng    | -0.02    |
| 21) Naphthalene-d8          | 7.95  | 136  | 1920915  | 20.00  | ng    | -0.02    |
| 38) Acenaphthene-d10        | 9.70  | 164  | 884955   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.18 | 188  | 1583918  | 20.00  | ng    | -0.01    |
| 75) Chrysene-d12            | 13.85 | 240  | 1242531  | 20.00  | ng    | -0.02    |
| 86) Perylene-d12            | 15.36 | 264  | 1019299  | 20.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.24  | 112  | 3170287  | 123.58 | ng    | -0.01    |
| 7) Phenol-d6                | 6.31  | 99   | 3745414  | 118.93 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.23  | 82   | 2300232  | 74.39  | ng    | -0.02    |
| 41) 2,4,6-Tribromophenol    | 10.49 | 330  | 934990   | 111.17 | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 9.04  | 172  | 4476851  | 88.91  | ng    | -0.01    |
| 78) Terphenyl-d14           | 12.78 | 244  | 3525352  | 64.19  | ng    | -0.01    |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF082316\  
Data File : BF089861.D  
Acq On : 24 Aug 2016 5:08  
Operator : UM/SJ  
Sample : BL-CHIPS-4  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
BL-CHIPS-4

Quant Time: Aug 25 14:32:29 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF081916.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Aug 19 14:02:57 2016  
Response via : Initial Calibration

