

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF082415\  
 Data File : BF081164.D  
 Acq On : 24 Aug 2015 12:17  
 Operator : UM/IZ  
 Sample : PB85134BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB85134BL

Quant Time: Aug 24 15:06:33 2015  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF081715.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 24 13:58:19 2015  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.54  | 152  | 99469    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.83  | 136  | 418172   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.57  | 164  | 193027   | 20.00 | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.04 | 188  | 414201   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12            | 13.65 | 240  | 315232   | 20.00 | ng    | -0.01    |
| 86) Perylene-d12            | 14.99 | 264  | 242896   | 20.00 | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.12  | 112  | 485608   | 73.43 | ng    | 0.02     |
| 7) Phenol-d6                | 6.20  | 99   | 721704   | 83.64 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.11  | 82   | 387253   | 42.30 | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.36 | 330  | 125209   | 74.83 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 8.90  | 172  | 622546   | 42.26 | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.62 | 244  | 612669   | 41.67 | ng    | -0.01    |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF082415\  
Data File : BF081164.D  
Acq On : 24 Aug 2015 12:17  
Operator : UM/IZ  
Sample : PB85134BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB85134BL

Quant Time: Aug 24 15:06:33 2015  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF081715.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Aug 24 13:58:19 2015  
Response via : Initial Calibration

