

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082415\
 Data File : BF081165.D
 Acq On : 24 Aug 2015 12:48
 Operator : UM/IZ
 Sample : G3083-12DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 8535DL

Quant Time: Aug 24 14:03:12 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 24 13:58:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.54	152	63127	20.00	ng	0.00
21) Naphthalene-d8	7.83	136	245171	20.00	ng	0.00
38) Acenaphthene-d10	9.57	164	116941	20.00	ng	-0.01
63) Phenanthrene-d10	11.04	188	225965	20.00	ng	0.00
75) Chrysene-d12	13.65	240	199187	20.00	ng	-0.01
86) Perylene-d12	14.99	264	144644	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.10	112	74783	17.82	ng	0.00
7) Phenol-d6	6.18	99	66010	12.05	ng	-0.01
23) Nitrobenzene-d5	7.11	82	97283	18.13	ng	-0.01
41) 2,4,6-Tribromophenol	10.34	330	26684	26.32	ng	-0.01
44) 2-Fluorobiphenyl	8.90	172	182921	20.50	ng	0.00
78) Terphenyl-d14	12.62	244	174763	18.81	ng	-0.01
Target Compounds						
8) 2-Chlorophenol	6.32	128	45140	9.83	ng	# 81
10) Phenol	6.20	94	82069	12.69	ng	# 64
17) 2-Methylphenol	6.81	107	27333	6.93	ng	# 91
20) 3+4-Methylphenols	6.96	107	46859	9.37	ng	# 45
26) 2-Nitrophenol	7.45	139	72402	31.02	ng	# 59
27) 2,4-Dimethylphenol	7.50	122	38713	9.38	ng	92
29) 2,4-Dichlorophenol	7.69	162	20171	5.80	ng	95
36) 4-Chloro-3-methylphenol	8.39	107	148653	35.88	ng	94
42) 2,4,6-Trichlorophenol	8.81	196	55505	20.82	ng	96
43) 2,4,5-Trichlorophenol	8.85	196	44162	16.58	ng	# 76
53) 2,4-Dinitrophenol	9.64	184	13542	16.16	ng	# 71
55) 4-Nitrophenol	9.69	139	20621	11.35	ng	# 86
69) Pentachlorophenol	10.85	266	20444	15.16	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082415\
Data File : BF081165.D
Acq On : 24 Aug 2015 12:48
Operator : UM/IZ
Sample : G3083-12DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
8535DL

Quant Time: Aug 24 14:03:12 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 24 13:58:19 2015
Response via : Initial Calibration

