

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082416\
 Data File : BF089866.D
 Acq On : 24 Aug 2016 11:36
 Operator : UM/SJ
 Sample : PB93047BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB93047BL

Quant Time: Aug 25 15:15:07 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 19 14:02:57 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	380170	20.00	ng	-0.02
21) Naphthalene-d8	7.95	136	1628356	20.00	ng	-0.02
38) Acenaphthene-d10	9.70	164	731128	20.00	ng	-0.01
63) Phenanthrene-d10	11.18	188	1337764	20.00	ng	-0.01
75) Chrysene-d12	13.90	240	1084602	20.00	ng	0.02
86) Perylene-d12	15.46	264	866548	20.00	ng	0.09
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	3295047	148.64	ng	-0.01
7) Phenol-d6	6.31	99	3563998	130.97	ng	-0.01
23) Nitrobenzene-d5	7.23	82	2018097	76.99	ng	-0.02
41) 2,4,6-Tribromophenol	10.49	330	763428	109.87	ng	-0.01
44) 2-Fluorobiphenyl	9.04	172	3927371	94.41	ng	-0.01
78) Terphenyl-d14	12.79	244	3334829	69.56	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082416\
Data File : BF089866.D
Acq On : 24 Aug 2016 11:36
Operator : UM/SJ
Sample : PB93047BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93047BL

Quant Time: Aug 25 15:15:07 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 19 14:02:57 2016
Response via : Initial Calibration

