

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082416\
 Data File : BF089872.D
 Acq On : 24 Aug 2016 14:15
 Operator : UM/SJ
 Sample : H4551-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AST-3

Quant Time: Aug 25 13:38:35 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 19 14:02:57 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	400178	20.00	ng	-0.02
21) Naphthalene-d8	7.95	136	1732959	20.00	ng	-0.02
38) Acenaphthene-d10	9.70	164	768516	20.00	ng	-0.01
63) Phenanthrene-d10	11.18	188	1368289	20.00	ng	-0.01
75) Chrysene-d12	13.83	240	1132280	20.00	ng	-0.05
86) Perylene-d12	15.29	264	920811	20.00	ng	-0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	3321363	142.33	ng	-0.01
7) Phenol-d6	6.31	99	3883442	135.57	ng	-0.01
23) Nitrobenzene-d5	7.23	82	2467963	88.47	ng	-0.02
41) 2,4,6-Tribromophenol	10.49	330	891956	122.13	ng	-0.01
44) 2-Fluorobiphenyl	9.04	172	4214928	96.39	ng	-0.01
78) Terphenyl-d14	12.78	244	3562197	71.17	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.43	163	1003386	18.54	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082416\
Data File : BF089872.D
Acq On : 24 Aug 2016 14:15
Operator : UM/SJ
Sample : H4551-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AST-3

Quant Time: Aug 25 13:38:35 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 19 14:02:57 2016
Response via : Initial Calibration

