

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082417\
 Data File : BF098012.D
 Acq On : 24 Aug 2017 20:02
 Operator : SJ/JU
 Sample : I4925-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RWB-2

Manual Integrations
 APPROVED

Sohil
 8/25/2017 5:40:12 PM

Quant Time: Aug 25 02:12:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 15:59:04 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	109012	20.00	ng	-0.01
21) Naphthalene-d8	8.02	136	432634	20.00	ng	-0.01
38) Acenaphthene-d10	9.77	164	188180	20.00	ng	-0.02
63) Phenanthrene-d10	11.25	188	361162	20.00	ng	-0.01
75) Chrysene-d12	13.88	240	272094	20.00	ng	-0.02
86) Perylene-d12	15.30	264	233616	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	973720	135.87	ng	0.00
7) Phenol-d6	6.37	99	1291947	132.68	ng	-0.01
23) Nitrobenzene-d5	7.30	82	751153	94.32	ng	-0.02
41) 2,4,6-Tribromophenol	10.56	330	214970	131.66	ng	-0.01
44) 2-Fluorobiphenyl	9.10	172	886544	69.22	ng	-0.01
78) Terphenyl-d14	12.84	244	734514	56.29	ng	-0.01
Target Compounds						
10) Phenol	6.39	94	31972	2.807	ng	91
49) Dimethylphthalate	9.49	163	171448	12.464	ng	99
74) Fluoranthene	12.46	202	88412	4.401	ng	98
77) Pyrene	12.69	202	80267	3.607	ng	99
80) Benzo(a)anthracene	13.87	228	40742	2.498	ng	96
82) Chrysene	13.90	228	41173	2.538	ng	96
87) Benzo(b)fluoranthene	14.89	252	46345m	3.147	ng	
89) Benzo(a)pyrene	15.23	252	32983	2.530	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082417\
Data File : BF098012.D
Acq On : 24 Aug 2017 20:02
Operator : SJ/JU
Sample : I4925-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RWB-2

Manual Integrations
APPROVED

Sohil
8/25/2017 5:40:12 PM

Quant Time: Aug 25 02:12:04 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration

