

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082621\
 Data File : BF125230.D
 Acq On : 26 Aug 2021 17:52
 Operator : JU/CG
 Sample : M3547-05
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PE-4

Quant Time: Aug 27 05:01:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	200538	20.000	ng	0.00
21) Naphthalene-d8	8.198	136	775507	20.000	ng	# 0.00
39) Acenaphthene-d10	9.951	164	412864	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	733949	20.000	ng	# 0.00
76) Chrysene-d12	14.074	240	476369	20.000	ng	0.00
86) Perylene-d12	15.568	264	430062	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	1192033	89.971	ng	0.00
7) Phenol-d6	6.534	99	1493345	87.150	ng	0.00
23) Nitrobenzene-d5	7.475	82	956960	62.124	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	441482	107.122	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	1649285	55.676	ng	0.00
79) Terphenyl-d14	13.021	244	1719572	57.507	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082621\
Data File : BF125230.D
Acq On : 26 Aug 2021 17:52
Operator : JU/CG
Sample : M3547-05
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PE-4

Quant Time: Aug 27 05:01:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 12:39:21 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 820
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52




Instrument :
BNA_F
ClientSampleId :
PE-4


Tgt Ion:152 Resp: 200538
Ion Ratio Lower Upper
152 100
150 159.4 117.6 176.4
115 64.3 47.1 70.7



#5
2-Fluorophenol
Concen: 89.971 ng
RT: 5.534 min Scan# 586
Delta R.T. 0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52

Tgt Ion:112 Resp: 1192033
Ion Ratio Lower Upper
112 100
64 70.0 44.2 66.2#
63 36.0 26.2 39.2









#7
Phenol-d6
Concen: 87.150 ng
RT: 6.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52




Instrument :
BNA_F
ClientSampleId :
PE-4


Tgt Ion: 99 Resp: 1493345
Ion Ratio Lower Upper
99 100
42 20.0 23.5 35.3#
71 35.2 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.198 min Scan# 1039
Delta R.T. -0.000 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52

Tgt Ion: 136 Resp: 775507
Ion Ratio Lower Upper
136 100
137 11.0 9.8 14.6
54 9.6 6.4 9.6
68 6.0 3.3 4.9#









#23
Nitrobenzene-d5
Concen: 62.124 ng
RT: 7.475 min Scan# 916
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52




Instrument :
BNA_F
ClientSampled :
PE-4


Tgt Ion: 82 Resp: 956960
Ion Ratio Lower Upper
82 100
128 40.9 40.6 60.8
54 60.1 49.6 74.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.951 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52

Tgt Ion:164 Resp: 412864
Ion Ratio Lower Upper
164 100
162 99.6 81.3 121.9
160 44.2 38.0 57.0









#42
2,4,6-Tribromophenol
Concen: 107.122 ng
RT: 10.739 min Scan# 1471
Delta R.T. -0.000 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52




Instrument :
BNA_F
ClientSampled :
PE-4


Tgt Ion:330 Resp: 441482
Ion Ratio Lower Upper
330 100
332 96.3 76.2 114.4
141 29.3 44.6 66.8#



#45
2-Fluorobiphenyl
Concen: 55.676 ng
RT: 9.269 min Scan# 1221
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52

Tgt Ion:172 Resp: 1649285
Ion Ratio Lower Upper
172 100
171 34.9 26.5 39.7
170 24.2 18.8 28.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1589
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52




Instrument :
BNA_F
ClientSampleId :
PE-4


Tgt Ion:188 Resp: 733949
Ion Ratio Lower Upper
188 100
94 10.7 6.7 10.1#
80 11.3 7.5 11.3#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2038
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52


Tgt Ion:240 Resp: 476369
Ion Ratio Lower Upper
240 100
120 11.5 10.8 16.2
236 26.5 21.6 32.4








#79
Terphenyl-d14
Concen: 57.507 ng
RT: 13.021 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52


Instrument :
BNA_F
ClientSampleId :
PE-4



Tgt Ion:244 Resp: 1719572
Ion Ratio Lower Upper
244 100
212 6.6 5.9 8.9
122 9.4 9.0 13.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BF125230.D
Acq: 26 Aug 2021 17:52



Tgt Ion:264 Resp: 430062
Ion Ratio Lower Upper
264 100
260 23.4 19.6 29.4
265 22.0 17.2 25.8

