

Method Path : Z:\HPCHEM1\BNA_F\METHODS\
 Method File : 625-TERPINEOL-BF082715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 27 06:31:06 2015
 Response Via : Initial Calibration

Calibration Files

2.5 =BF081235.D 10 =BF081236.D 25 =BF081237.D 40 =BF081238.D 60 =BF081240.D 50 =BF081239.D 80 =BF081241.D

	Compound	2.5	10	25	40	60	50	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2) S	2-Fluorophenol	1.276	1.191	1.297	1.188	1.271	1.353	1.263	1.263	4.62
3) S	Phenol-d6	1.803	1.565	1.909	1.506	1.747	1.592	1.682	1.686	8.49
4) I	Naphthalene-d8	-----ISTD-----								
5) S	Nitrobenzene-d5	0.415	0.378	0.471	0.381	0.386	0.370	0.456	0.408	9.94
6)	alpha terpinol	0.030	0.039	0.043	0.040	0.043	0.043	0.042	0.040	11.96
7) I	Acenaphthene-d10	-----ISTD-----								
8) S	2,4,6-Tribromo...	0.121	0.135	0.148	0.142	0.163	0.144	0.167	0.146	10.84
9) S	2-Fluorobiphenyl	1.581	1.358	1.473	1.195	1.141	1.186	1.221	1.308	12.78
10) I	Phenanthrene-d10	-----ISTD-----								
11) I	Chrysene-d12	-----ISTD-----								
12) S	Terphenyl-d14	0.971	0.812	0.942	0.801	0.717	0.822	0.845	0.844	10.31
13) I	Perylene-d12	-----ISTD-----								

(#) = Out of Range