

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082715\
 Data File : BF081242.D
 Acq On : 27 Aug 2015 6:29
 Operator : UM/IZ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF082715

Quant Time: Aug 27 07:08:10 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\625-TERPINEOL-BF082715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 06:31:06 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2 S	2-Fluorophenol	80.000	76.079	4.9	101	0.00
3 S	Phenol-d6	80.000	70.652	11.7	99	0.00
4 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
5 S	Nitrobenzene-d5	80.000	76.312	4.6	101	0.00
6	alpha terpinol	40.000	40.436	-1.1	101	0.00
7 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
8 S	2,4,6-Tribromophenol	80.000	76.564	4.3	99	0.00
9 S	2-Fluorobiphenyl	80.000	75.361	5.8	104	0.00
10 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
11 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
12 S	Terphenyl-d14	80.000	82.053	-2.6	105	0.00
13 I	Perylene-d12	20.000	20.000	0.0	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0