

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF082715\  
 Data File : BF081245.D  
 Acq On : 27 Aug 2015 11:13  
 Operator : UM/IZ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: Aug 27 11:46:44 2015  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\625-TERPINEOL-BF082715.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 27 06:31:06 2015  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0  | 97    | 0.00     |
| 2 S  | 2-Fluorophenol         | 80.000 | 74.730 | 6.6  | 97    | 0.00     |
| 3 S  | Phenol-d6              | 80.000 | 75.579 | 5.5  | 103   | 0.00     |
| 4 I  | Naphthalene-d8         | 20.000 | 20.000 | 0.0  | 101   | 0.00     |
| 5 S  | Nitrobenzene-d5        | 80.000 | 73.284 | 8.4  | 99    | 0.00     |
| 6    | alpha terpinol         | 40.000 | 38.942 | 2.6  | 100   | 0.00     |
| 7 I  | Acenaphthene-d10       | 20.000 | 20.000 | 0.0  | 97    | 0.00     |
| 8 S  | 2,4,6-Tribromophenol   | 80.000 | 77.816 | 2.7  | 97    | 0.00     |
| 9 S  | 2-Fluorobiphenyl       | 80.000 | 75.599 | 5.5  | 101   | 0.00     |
| 10 I | Phenanthrene-d10       | 20.000 | 20.000 | 0.0  | 104   | 0.00     |
| 11 I | Chrysene-d12           | 20.000 | 20.000 | 0.0  | 103   | 0.00     |
| 12 S | Terphenyl-d14          | 80.000 | 74.642 | 6.7  | 101   | 0.00     |
| 13 I | Perylene-d12           | 20.000 | 20.000 | 0.0  | 102   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0