

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082721\
 Data File : BF125246.D
 Acq On : 27 Aug 2021 9:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 12:28:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.652	0.619	5.1	108	0.00
3	Pyridine	1.721	1.521	11.6	98	0.00
4	n-Nitrosodimethylamine	0.861	0.694	19.4	88	0.00
5 S	2-Fluorophenol	1.321	1.197	9.4	106	0.00
6	Aniline	2.007	1.772	11.7	99	0.00
7 S	Phenol-d6	1.709	1.507	11.8	101	0.00
8	2-Chlorophenol	1.348	1.222	9.3	104	0.00
9	Benzaldehyde	1.199	0.957	20.2	96	0.00
10 C	Phenol	1.790	1.690	5.6	108	0.00
11	bis(2-Chloroethyl)ether	1.409	1.244	11.7	102	0.00
12	1,3-Dichlorobenzene	1.562	1.413	9.5	106	0.00
13 C	1,4-Dichlorobenzene	1.556	1.430	8.1	107	0.00
14	1,2-Dichlorobenzene	1.464	1.324	9.6	106	0.00
15	Benzyl Alcohol	1.316	1.088	17.3	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.827	2.316	18.1	92	0.00
17	2-Methylphenol	1.132	1.025	9.5	102	0.00
18	Hexachloroethane	0.545	0.495	9.2	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.028	0.874	15.0	94	0.00
20	3+4-Methylphenols	1.421	1.232	13.3	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
22	Acetophenone	0.515	0.448	13.0	98	0.00
23 S	Nitrobenzene-d5	0.397	0.367	7.6	99	0.00
24	Nitrobenzene	0.433	0.394	9.0	98	0.00
25	Isophorone	0.766	0.667	12.9	91	0.00
26 C	2-Nitrophenol	0.173	0.177	-2.3	102	0.00
27	2,4-Dimethylphenol	0.303	0.255	15.8	91	0.00
28	bis(2-Chloroethoxy)methane	0.424	0.377	11.1	97	0.00
29 C	2,4-Dichlorophenol	0.309	0.284	8.1	98	0.00
30	1,2,4-Trichlorobenzene	0.337	0.306	9.2	101	0.00
31	Naphthalene	1.024	0.895	12.6	98	0.00
32	Benzoic acid	0.205	0.196	4.4	86	0.00
33	4-Chloroaniline	0.404	0.362	10.4	95	0.00
34 C	Hexachlorobutadiene	0.215	0.200	7.0	106	0.00
35	Caprolactam	0.080	0.073	8.8	82	0.00
36 C	4-Chloro-3-methylphenol	0.315	0.286	9.2	93	0.00
37	2-Methylnaphthalene	0.677	0.602	11.1	97	0.00
38	1-Methylnaphthalene	0.632	0.563	10.9	96	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.653	0.613	6.1	101	0.00
41 P	Hexachlorocyclopentadiene	0.343	0.279	18.7	85	0.00
42 S	2,4,6-Tribromophenol	0.200	0.207	-3.5	95	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.428	7.0	95	0.00
44	2,4,5-Trichlorophenol	0.484	0.456	5.8	94	0.00
45 S	2-Fluorobiphenyl	1.435	1.274	11.2	100	0.00
46	1,1'-Biphenyl	1.612	1.413	12.3	94	0.00
47	2-Chloronaphthalene	1.310	1.168	10.8	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082721\
 Data File : BF125246.D
 Acq On : 27 Aug 2021 9:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 12:28:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.406	0.382	5.9	84	0.00
49	Acenaphthylene	1.909	1.683	11.8	90	0.00
50	Dimethylphthalate	1.386	1.250	9.8	89	0.00
51	2,6-Dinitrotoluene	0.301	0.283	6.0	86	0.00
52 C	Acenaphthene	1.184	1.083	8.5	93	0.00
53	3-Nitroaniline	0.324	0.298	8.0	83	0.00
54 P	2,4-Dinitrophenol	0.119	0.135	-13.4	91	0.00
55	Dibenzofuran	1.703	1.509	11.4	91	0.00
56 P	4-Nitrophenol	0.257	0.233	9.3	78	0.00
57	2,4-Dinitrotoluene	0.364	0.368	-1.1	88	0.00
58	Fluorene	1.259	1.135	9.8	93	0.00
59	2,3,4,6-Tetrachlorophenol	0.359	0.342	4.7	91	0.00
60	Diethylphthalate	1.351	1.211	10.4	86	0.00
61	4-Chlorophenyl-phenylether	0.637	0.589	7.5	95	0.00
62	4-Nitroaniline	0.291	0.281	3.4	81	0.00
63	Azobenzene	1.516	1.296	14.5	84	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.118	-11.3	90	0.00
66 c	n-Nitrosodiphenylamine	0.716	0.646	9.8	88	0.00
67	4-Bromophenyl-phenylether	0.240	0.230	4.2	90	0.00
68	Hexachlorobenzene	0.247	0.235	4.9	89	0.00
69	Atrazine	0.205	0.192	6.3	83	0.00
70 C	Pentachlorophenol	0.144	0.136	5.6	82	0.00
71	Phenanthrene	1.102	0.973	11.7	84	0.00
72	Anthracene	1.086	0.982	9.6	86	0.00
73	Carbazole	0.993	0.896	9.8	83	0.00
74	Di-n-butylphthalate	1.178	1.100	6.6	85	0.00
75 C	Fluoranthene	1.093	1.010	7.6	84	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
77	Benzydine	0.698	0.551	21.1	79	0.00
78	Pyrene	1.716	1.448	15.6	86	0.00
79 S	Terphenyl-d14	1.255	1.069	14.8	90	0.00
80	Butylbenzylphthalate	0.674	0.610	9.5	87	0.00
81	Benzo(a)anthracene	1.441	1.296	10.1	98	0.00
82	3,3'-Dichlorobenzidine	0.447	0.425	4.9	101	0.00
83	Chrysene	1.425	1.263	11.4	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.927	0.860	7.2	96	0.00
85 c	Di-n-octyl phthalate	1.541	1.461	5.2	102	0.00
86 I	Perylene-d12	1.000	1.000	0.0	146	0.00
87	Indeno(1,2,3-cd)pyrene	1.441	1.445	-0.3	158#	0.00
88	Benzo(b)fluoranthene	1.461	1.336	8.6	123	0.00
89	Benzo(k)fluoranthene	1.363	1.152	15.5	113	0.00
90 C	Benzo(a)pyrene	1.291	1.177	8.8	127	0.00
91	Dibenzo(a,h)anthracene	1.245	1.233	1.0	154#	0.00
92	Benzo(g,h,i)perylene	1.201	1.252	-4.2	166#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082721\
Data File : BF125246.D
Acq On : 27 Aug 2021 9:43
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 27 12:28:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 12:39:21 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0