

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082724\
 Data File : BF139175.D
 Acq On : 27 Aug 2024 09:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 11:14:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 02:51:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	41.833	-4.6	97	0.00
3	Pyridine	40.000	41.463	-3.7	95	0.00
4	n-Nitrosodimethylamine	40.000	40.463	-1.2	92	0.00
5 S	2-Fluorophenol	80.000	85.597	-7.0	94	0.00
6	Aniline	40.000	40.825	-2.1	93	-0.03
7 S	Phenol-d6	80.000	81.840	-2.3	93	0.00
8	2-Chlorophenol	40.000	44.445	-11.1	96	0.00
9	Benzaldehyde	40.000	40.027	-0.1	94	0.00
10 C	Phenol	40.000	41.715	-4.3	93	0.00
11	bis(2-Chloroethyl)ether	40.000	39.327	1.7	94	0.00
12	1,3-Dichlorobenzene	40.000	39.770	0.6	91	0.00
13 C	1,4-Dichlorobenzene	40.000	40.230	-0.6	93	0.00
14	1,2-Dichlorobenzene	40.000	39.628	0.9	92	0.00
15	Benzyl Alcohol	40.000	42.647	-6.6	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.167	2.1	90	0.00
17	2-Methylphenol	40.000	41.130	-2.8	94	0.00
18	Hexachloroethane	40.000	42.935	-7.3	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.485	-3.7	92	0.00
20	3+4-Methylphenols	40.000	41.877	-4.7	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	41.432	-3.6	92	0.00
23 S	Nitrobenzene-d5	80.000	94.312	-17.9	99	0.00
24	Nitrobenzene	40.000	43.638	-9.1	97	0.00
25	Isophorone	40.000	42.110	-5.3	92	0.00
26 C	2-Nitrophenol	40.000	45.751	-14.4	112	0.00
27	2,4-Dimethylphenol	40.000	44.841	-12.1	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.990	-2.5	90	0.00
29 C	2,4-Dichlorophenol	40.000	46.097	-15.2	92	0.00
30	1,2,4-Trichlorobenzene	40.000	40.945	-2.4	91	0.00
31	Naphthalene	40.000	41.245	-3.1	91	0.00
32	Benzoic acid	40.000	43.591	-9.0	108	0.00
33	4-Chloroaniline	40.000	41.356	-3.4	92	0.00
34 C	Hexachlorobutadiene	40.000	41.608	-4.0	90	0.00
35	Caprolactam	40.000	46.313	-15.8	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.994	-12.5	94	0.00
37	2-Methylnaphthalene	40.000	41.285	-3.2	91	0.00
38	1-Methylnaphthalene	40.000	41.331	-3.3	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.682	-1.7	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.296	-8.2	92	0.00
42 S	2,4,6-Tribromophenol	80.000	91.634	-14.5	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	48.182	-20.5#	96	0.00
44	2,4,5-Trichlorophenol	40.000	46.352	-15.9	94	0.00
45 S	2-Fluorobiphenyl	80.000	79.745	0.3	91	0.00
46	1,1'-Biphenyl	40.000	41.077	-2.7	92	0.00
47	2-Chloronaphthalene	40.000	40.702	-1.8	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082724\
 Data File : BF139175.D
 Acq On : 27 Aug 2024 09:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 27 11:14:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 02:51:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.438	-8.6	104	0.00
49	Acenaphthylene	40.000	41.415	-3.5	93	0.00
50	Dimethylphthalate	40.000	43.774	-9.4	94	0.00
51	2,6-Dinitrotoluene	40.000	44.566	-11.4	103	0.00
52 C	Acenaphthene	40.000	41.462	-3.7	93	0.00
53	3-Nitroaniline	40.000	44.485	-11.2	103	0.00
54 P	2,4-Dinitrophenol	40.000	45.355	-13.4	114	0.00
55	Dibenzofuran	40.000	41.232	-3.1	93	0.00
56 P	4-Nitrophenol	40.000	44.360	-10.9	102	0.00
57	2,4-Dinitrotoluene	40.000	45.042	-12.6	109	0.00
58	Fluorene	40.000	41.494	-3.7	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	46.952	-17.4	94	0.00
60	Diethylphthalate	40.000	44.358	-10.9	94	0.00
61	4-Chlorophenyl-phenylether	40.000	41.138	-2.8	93	0.00
62	4-Nitroaniline	40.000	46.041	-15.1	102	0.00
63	Azobenzene	40.000	41.914	-4.8	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.251	-13.1	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.147	-2.9	94	0.00
67	4-Bromophenyl-phenylether	40.000	40.745	-1.9	94	0.00
68	Hexachlorobenzene	40.000	41.766	-4.4	96	0.00
69	Atrazine	40.000	44.951	-12.4	99	0.00
70 C	Pentachlorophenol	40.000	46.582	-16.5	102	0.00
71	Phenanthrene	40.000	40.844	-2.1	94	0.00
72	Anthracene	40.000	41.306	-3.3	95	0.00
73	Carbazole	40.000	43.132	-7.8	98	0.00
74	Di-n-butylphthalate	40.000	50.494	-26.2#	103	0.00
75 C	Fluoranthene	40.000	44.554	-11.4	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	60.207	-50.5#	144	0.00
78	Pyrene	40.000	41.515	-3.8	101	0.00
79 S	Terphenyl-d14	80.000	83.746	-4.7	105	0.00
80	Butylbenzylphthalate	40.000	47.671	-19.2	125	0.00
81	Benzo(a)anthracene	40.000	42.005	-5.0	108	0.00
82	3,3'-Dichlorobenzidine	40.000	44.336	-10.8	110	0.00
83	Chrysene	40.000	42.256	-5.6	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.491	-3.7	102	0.00
85 c	Di-n-octyl phthalate	40.000	33.838	15.4	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.017	-0.0	89	0.00
88	Benzo(b)fluoranthene	40.000	42.947	-7.4	98	0.00
89	Benzo(k)fluoranthene	40.000	42.819	-7.0	97	0.00
90 C	Benzo(a)pyrene	40.000	42.597	-6.5	94	0.00
91	Dibenzo(a,h)anthracene	40.000	40.236	-0.6	90	0.00
92	Benzo(g,h,i)perylene	40.000	39.008	2.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082724\
Data File : BF139175.D
Acq On : 27 Aug 2024 09:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 27 11:14:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 27 02:51:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1