




Data Path : Z:\HPCHEM1\BNA_F\Data\BF082815\
Data File : BF081246.D
Acq On : 27 Aug 2015 13:56
Operator : UM/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 28 00:27:42 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 24 13:58:19 2015
Response via : Initial Calibration



TIC: BF081246.D

(69) Pentachlorophenol (C)




10.812min (-0.047) 0.00ng

response 44358

Ion	Exp%	Act%
265.70	100	100
268.00	64.00	61.21
264.00	63.20	63.98
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF082815\
Data File : BF081246.D
Acq On : 27 Aug 2015 13:56
Operator : UM/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 28 00:27:42 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 24 13:58:19 2015
Response via : Initial Calibration



TIC: BF081246.D

(76) Benzidine

12.332min (-0.047) 0.00ng

response 256921

Ion	Exp%	Act%
184.00	100	100
185.00	15.00	14.00
183.00	11.10	10.49
0.00	0.00	0.00