

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082817\
 Data File : BF098102.D
 Acq On : 29 Aug 2017 8:25
 Operator : SJ/JU
 Sample : I4961-10
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-02-COMP

Quant Time: Aug 29 09:37:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 15:59:04 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	113471	20.00	ng	-0.01
21) Naphthalene-d8	8.02	136	449585	20.00	ng	-0.01
38) Acenaphthene-d10	9.77	164	194935	20.00	ng	-0.01
63) Phenanthrene-d10	11.26	188	365003	20.00	ng	0.00
75) Chrysene-d12	13.89	240	282668	20.00	ng	-0.01
86) Perylene-d12	15.31	264	277398	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	945809	126.79	ng	0.00
7) Phenol-d6	6.38	99	1287408	127.01	ng	0.00
23) Nitrobenzene-d5	7.30	82	800391	96.71	ng	-0.02
41) 2,4,6-Tribromophenol	10.56	330	219205	129.60	ng	-0.01
44) 2-Fluorobiphenyl	9.10	172	1108246	83.53	ng	-0.01
78) Terphenyl-d14	12.84	244	851638	62.83	ng	0.00
Target Compounds						
10) Phenol	6.39	94	44594	3.762	ng	Qvalue 90
49) Dimethylphthalate	9.49	163	172356	12.096	ng	99
70) Phenanthrene	11.28	178	105432	5.421	ng	96
74) Fluoranthene	12.46	202	92222	4.543	ng	98
77) Pyrene	12.69	202	87025	3.764	ng	97
80) Benzo(a)anthracene	13.87	228	41842	2.470	ng	97
82) Chrysene	13.91	228	41702	2.474	ng	96
87) Benzo(b)fluoranthene	14.90	252	45379m	2.595	ng	
89) Benzo(a)pyrene	15.24	252	33903	2.190	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF082817\
Data File : BF098102.D
Acq On : 29 Aug 2017 8:25
Operator : SJ/JU
Sample : I4961-10
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-02-COMP

Quant Time: Aug 29 09:37:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration

