

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082820\
 Data File : BF121471.D
 Acq On : 28 Aug 2020 21:05
 Operator : JU/CG
 Sample : L3837-10
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 789UMR-TP06-0006-01

Quant Time: Aug 29 00:36:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 28 09:25:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	311855	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	1225016	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	627814	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	1164855	20.00	ng	0.00
76) Chrysene-d12	14.02	240	764257	20.00	ng	0.00
86) Perylene-d12	15.49	264	717416	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.47	112	1530031	82.31	ng	0.00
7) Phenol-d6	6.48	99	1997811	76.91	ng	-0.01
23) Nitrobenzene-d5	7.42	82	1273630	71.08	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	576322	90.88	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	2248732	57.79	ng	0.00
79) Terphenyl-d14	12.97	244	2117323	58.32	ng	0.00

Target Compounds


						Qvalue
9) Benzaldehyde	6.40	77	1482	Below Cal	#	83
19) n-Nitroso-di-n-propylamine	7.42	70	170195	11.724 ng	#	77
25) Isophorone	7.42	82	1272829	31.581 ng	#	64
32) Benzoic acid	7.82	122	344	6.702 ng	#	62
48) 2-Nitroaniline	9.42	65	247	6.667 ng	#	5
50) Dimethylphthalate	9.61	163	203173	4.711 ng		98
51) 2,6-Dinitrotoluene	9.90	165	83026	15.318 ng	#	24
56) 4-Nitrophenol	10.10	139	4131	8.304 ng	#	6
57) 2,4-Dinitrotoluene	9.90	165	83026	14.356 ng	#	21
65) 4,6-Dinitro-2-methylphenol	10.68	198	1166	5.994 ng	#	1
67) 4-Bromophenyl-phenylether	10.69	248	34212	2.670 ng	#	1
71) Phenanthrene	11.40	178	338632	5.665 ng		97
75) Fluoranthene	12.59	202	814291	12.639 ng		97
78) Pyrene	12.82	202	659957	12.091 ng		96
81) Benzo(a)anthracene	14.01	228	270547	5.799 ng		98
83) Chrysene	14.04	228	313563	6.718 ng		96
87) Indeno(1,2,3-cd)pyrene	16.94	276	194747	4.412 ng		94
88) Benzo(b)fluoranthene	15.06	252	534322	11.450 ng		99
89) Benzo(k)fluoranthene	15.06	252	534322	12.310 ng		99
90) Benzo(a)pyrene	15.42	252	258143	6.152 ng		100
92) Benzo(g,h,i)perylene	17.38	276	185846	5.339 ng		100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082820\
Data File : BF121471.D
Acq On : 28 Aug 2020 21:05
Operator : JU/CG
Sample : L3837-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01

Quant Time: Aug 29 00:36:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 28 09:25:49 2020
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.86 min Scan# 811
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.7	125.8	188.6
115	61.2	46.6	69.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5




2-Fluorophenol
Concen: 82.314 ng
RT: 5.47 min Scan# 576
Delta R.T. -0.00 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion	Ratio	Lower	Upper
112	100		
64	61.4	48.2	72.2
63	31.5	24.8	37.2



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 76.910 ng

RT: 6.48 min Scan# 747

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01





Tot Ion: 99 Resp: 1997811

Ion Ratio Lower Upper

99 100

42 17.5 13.5 20.3

71 31.8 24.6 37.0



#19

n-Nitroso-di-n-propylamine

Concen: 11.724 ng

RT: 7.42 min Scan# 907

Delta R.T. 0.14 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Tgt Ion: 70 Resp: 170195


Ion Ratio Lower Upper


70 100

42 45.3 45.8 68.6#

101 0.0 7.5 11.3#

130 2.2 16.5 24.7#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.14 min Scan# 1029

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01

Tot Ion:136 Resp: 1225016


Ion Ratio Lower Upper

136 100

137 11.8 9.1 13.7

54 8.9 7.0 10.4

68 7.0 5.5 8.3






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#23

Nitrobenzene-d5

Concen: 71.079 ng

RT: 7.42 min Scan# 907

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion: 82 Resp: 1273630

Ion Ratio Lower Upper

82 100

128 47.0 34.1 51.1



54 50.0 41.0 61.6




Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF1





#25

Isophorone

Concen: 31.581 ng

RT: 7.42 min Scan# 907

Delta R.T. -0.26 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion: 82 Resp: 1272829

Ion Ratio Lower Upper

82 100

95 0.0 5.6 8.4#




138 0.0 14.8 22.2#



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#32

Benzoic acid

Concen: 6.702 ng

RT: 7.82 min Scan# 974

Delta R.T. -0.11 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion: 122 Resp: 344

Ion Ratio Lower Upper

122 100

105 141.5 100.8 140.8#



77 151.4 71.5 111.5#




Abundance Ion 122.00 (121.70 to 122.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 77.00 (76.70 to 77.70): BF1





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.90 min Scan# 1328

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:164 Resp: 627814

Ion Ratio Lower Upper

164 100

162 101.5 82.4 123.6




160 46.4 37.6 56.4



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 90.885 ng

RT: 10.69 min Scan# 1462

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:330 Resp: 576322

Ion Ratio Lower Upper

330 100

332 95.7 76.2 114.2



141 30.4 24.6 36.8




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#45
2-Fluorobiphenyl
Concen: 57.794 ng
RT: 9.22 min Scan# 1213
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05




Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01


Tot Ion:172 Resp: 2248732
Ion Ratio Lower Upper
172 100
171 37.2 30.6 45.8
170 25.3 20.6 31.0



#48
2-Nitroaniline
Concen: 6.667 ng
RT: 9.42 min Scan# 1246
Delta R.T. -0.02 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion: 65 Resp: 247
Ion Ratio Lower Upper
65 100
92 0.0 55.6 83.4#
138 0.0 82.9 124.3#





#50

Dimethylphthalate

Concen: 4.711 ng

RT: 9.61 min Scan# 1279

Delta R.T. -0.02 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:163 Resp: 203173

Ion Ratio Lower Upper




163 100

194 3.9 3.3 4.9

164 9.9 8.6 12.8



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#51

2,6-Dinitrotoluene

Concen: 15.318 ng

RT: 9.90 min Scan# 1328

Delta R.T. 0.21 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:165 Resp: 83026

Ion Ratio Lower Upper



165 100


63 1.1 43.6 65.4#

89 1.2 47.4 71.2#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1









#56
4-Nitrophenol
Concen: 8.304 ng
RT: 10.10 min Scan# 1362
Delta R.T. 0.10 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05




Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01


Tot Ion	Ratio	Lower	Upper
139	100		
109	0.0	50.5	90.5#
65	0.0	80.2	120.2#



#57
2,4-Dinitrotoluene
Concen: 14.356 ng
RT: 9.90 min Scan# 1328
Delta R.T. -0.19 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion	Ratio	Lower	Upper
165	100		
63	1.1	36.0	54.0#
89	1.2	62.6	93.8#
182	0.0	2.2	3.2#





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.38 min Scan# 1580

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:188 Resp: 1164855

Ion Ratio Lower Upper

188 100

94 10.9 8.3 12.5




80 11.6 8.6 13.0



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#65

4,6-Dinitro-2-methylphenol

Concen: 5.994 ng

RT: 10.68 min Scan# 1461

Delta R.T. 0.19 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:198 Resp: 1166

Ion Ratio Lower Upper

198 100

51 216.7 29.0 69.0#



105 211.6 29.1 69.1#




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 2.670 ng
 RT: 10.69 min Scan# 1462
 Delta R.T. -0.25 min
 Lab File: BF121471.D
 Acq: 28 Aug 2020 21:05

Instrument :
 BNA_F
 ClientSampleId :
 789UMR-TP06-0006-01

Tot Ion:248 Resp: 34212
 Ion Ratio Lower Upper
 248 100
 250 201.2 78.7 118.1#
 141 421.4 52.6 78.8#




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#71
 Phenanthrene
 Concen: 5.665 ng
 RT: 11.40 min Scan# 1584
 Delta R.T. -0.01 min
 Lab File: BF121471.D
 Acq: 28 Aug 2020 21:05

Tgt Ion:178 Resp: 338632
 Ion Ratio Lower Upper
 178 100
 176 19.3 16.9 25.3
 179 15.8 13.4 20.0



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#75

Fluoranthene

Concen: 12.639 ng

RT: 12.59 min Scan# 1786

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:202 Resp: 814291

Ion Ratio Lower Upper

202 100

101 12.0 0.0 33.3




203 17.7 0.0 39.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.02 min Scan# 2029

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:240 Resp: 764257

Ion Ratio Lower Upper

240 100

120 11.2 9.5 14.3



236 26.4 21.8 32.6




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Pvrene

Concen: 12.091 ng

RT: 12.82 min Scan# 1825

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:202 Resp: 659957

Ion Ratio Lower Upper

202 100

200 20.2 18.0 27.0




203 17.8 15.3 22.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#79

Terphenyl-d14

Concen: 58.316 ng

RT: 12.97 min Scan# 1850

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:244 Resp: 2117323

Ion Ratio Lower Upper

244 100

212 7.6 6.3 9.5



122 13.3 9.5 14.3




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#81
Benzo(a)anthracene
Concen: 5.799 ng
RT: 14.01 min Scan# 2027
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05




Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01


Tot Ion	Ratio	Lower	Upper
228	100		
226	30.9	23.0	34.4
229	20.7	16.6	24.8



#83
Chrysene
Concen: 6.718 ng
RT: 14.04 min Scan# 2033
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.3	26.2	39.2
229	21.6	16.6	25.0









#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.49 min Scan# 2278
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05




Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01


Tot Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.6	29.4
265	22.7	17.8	26.6



#87
Indeno(1,2,3-cd)pyrene
Concen: 4.412 ng
RT: 16.94 min Scan# 2525
Delta R.T. -0.02 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.0	20.6	31.0
277	24.5	20.5	30.7









#88
Benzo(b)fluoranthene
Concen: 11.450 ng
RT: 15.06 min Scan# 2205
Delta R.T. -0.01 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05




Instrument :
BNA_F
ClientSampleId :
789UMR-TP06-0006-01


Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.3	27.5
125	10.9	9.2	13.8



#89
Benzo(k)fluoranthene
Concen: 12.310 ng
RT: 15.06 min Scan# 2205
Delta R.T. -0.04 min
Lab File: BF121471.D
Acq: 28 Aug 2020 21:05

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.3	27.5
125	10.9	8.5	12.7





#90

Benzo(a)pyrene

Concen: 6.152 ng

RT: 15.42 min Scan# 2267

Delta R.T. -0.01 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05

Instrument :

BNA_F

ClientSampleId :

789UMR-TP06-0006-01


Tot Ion:252 Resp: 258143

Ion Ratio Lower Upper

252 100

253 22.7 18.0 27.0




125 11.1 9.0 13.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#92

Benzo(g,h,i)perylene

Concen: 5.339 ng

RT: 17.38 min Scan# 2600

Delta R.T. -0.02 min

Lab File: BF121471.D

Acq: 28 Aug 2020 21:05


Tgt Ion:276 Resp: 185846

Ion Ratio Lower Upper

276 100

277 23.2 18.9 28.3

138 21.7 17.4 26.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

