


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF082915\
 Data File : BF081268.D
 Acq On : 29 Aug 2015 11:45
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 31 13:41:31 2015



AutoFind: Scans 833, 834, 835; Background Corrected with Scan 830

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.2	9397	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	48.8	8017	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	56.5	9283	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	16417	PASS
199	198	5	9	5.7	943	PASS
275	198	10	60	19.2	3154	PASS
365	198	1	100	2.0	333	PASS
441	442	0.01	24	15.8	1376	PASS
442	198	50	100	53.0	8699	PASS
443	442	15	24	19.3	1681	PASS