

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082922\
 Data File : BF130032.D
 Acq On : 29 Aug 2022 13:37
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 30 01:25:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 30 01:22:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
2	1,4-Dioxane	0.498	0.510	-2.4	78	0.00
3	Pyridine	1.304	1.280	1.8	74	0.00
4	n-Nitrosodimethylamine	0.601	0.677	-12.6	83	0.00
5 S	2-Fluorophenol	1.208	1.260	-4.3	81	0.00
6	Aniline	1.724	1.807	-4.8	81	0.00
7 S	Phenol-d6	1.513	1.542	-1.9	79	0.00
8	2-Chlorophenol	1.345	1.390	-3.3	79	0.00
9	Benzaldehyde	0.893	0.906	-1.5	80	0.00
10 C	Phenol	1.659	1.729	-4.2	81	0.00
11	bis(2-Chloroethyl)ether	1.260	1.270	-0.8	77	0.00
12	1,3-Dichlorobenzene	1.452	1.475	-1.6	78	0.00
13 C	1,4-Dichlorobenzene	1.483	1.505	-1.5	79	0.00
14	1,2-Dichlorobenzene	1.378	1.408	-2.2	79	0.00
15	Benzyl Alcohol	1.142	1.171	-2.5	79	0.00
16	2,2'-oxybis(1-Chloropropane	1.688	1.738	-3.0	79	0.00
17	2-Methylphenol	1.080	1.116	-3.3	80	0.00
18	Hexachloroethane	0.557	0.587	-5.4	81	0.00
19 P	n-Nitroso-di-n-propylamine	0.956	1.014	-6.1	81	0.00
20	3+4-Methylphenols	1.449	1.546	-6.7	82	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
22	Acetophenone	0.487	0.509	-4.5	83	0.00
23 S	Nitrobenzene-d5	0.374	0.388	-3.7	81	0.00
24	Nitrobenzene	0.377	0.393	-4.2	82	0.00
25	Isophorone	0.643	0.657	-2.2	80	0.00
26 C	2-Nitrophenol	0.185	0.193	-4.3	80	0.00
27	2,4-Dimethylphenol	0.266	0.267	-0.4	78	0.00
28	bis(2-Chloroethoxy)methane	0.375	0.379	-1.1	77	0.00
29 C	2,4-Dichlorophenol	0.291	0.295	-1.4	78	0.00
30	1,2,4-Trichlorobenzene	0.307	0.313	-2.0	80	0.00
31	Naphthalene	1.046	1.061	-1.4	80	0.00
32	Benzoic acid	0.113	0.115	-1.8	76	0.00
33	4-Chloroaniline	0.411	0.428	-4.1	81	0.00
34 C	Hexachlorobutadiene	0.187	0.191	-2.1	79	0.00
35	Caprolactam	0.090	0.095	-5.6	82	0.00
36 C	4-Chloro-3-methylphenol	0.313	0.330	-5.4	83	0.00
37	2-Methylnaphthalene	0.686	0.691	-0.7	79	0.00
38	1-Methylnaphthalene	0.667	0.670	-0.4	79	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
40	1,2,4,5-Tetrachlorobenzene	0.556	0.554	0.4	79	0.00
41 P	Hexachlorocyclopentadiene	0.117	0.123	-5.1	75	0.00
42 S	2,4,6-Tribromophenol	0.201	0.195	3.0	76	0.00
43 C	2,4,6-Trichlorophenol	0.361	0.341	5.5	74	0.00
44	2,4,5-Trichlorophenol	0.389	0.370	4.9	73	0.00
45 S	2-Fluorobiphenyl	1.317	1.321	-0.3	81	0.00
46	1,1'-Biphenyl	1.519	1.504	1.0	79	0.00
47	2-Chloronaphthalene	1.202	1.210	-0.7	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082922\
 Data File : BF130032.D
 Acq On : 29 Aug 2022 13:37
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 30 01:25:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 30 01:22:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.380	0.409	-7.6	84	0.00
49	Acenaphthylene	1.823	1.820	0.2	79	0.00
50	Dimethylphthalate	1.434	1.462	-2.0	81	0.00
51	2,6-Dinitrotoluene	0.311	0.322	-3.5	80	0.00
52 C	Acenaphthene	1.106	1.128	-2.0	81	0.00
53	3-Nitroaniline	0.344	0.359	-4.4	82	0.00
54 P	2,4-Dinitrophenol	0.118	0.126	-6.8	75	0.00
55	Dibenzofuran	1.662	1.677	-0.9	82	0.00
56 P	4-Nitrophenol	0.153	0.159	-3.9	77	0.00
57	2,4-Dinitrotoluene	0.404	0.425	-5.2	85	0.00
58	Fluorene	1.387	1.379	0.6	80	0.00
59	2,3,4,6-Tetrachlorophenol	0.285	0.297	-4.2	79	0.00
60	Diethylphthalate	1.430	1.434	-0.3	80	0.00
61	4-Chlorophenyl-phenylether	0.623	0.622	0.2	80	0.00
62	4-Nitroaniline	0.347	0.357	-2.9	82	0.00
63	Azobenzene	1.366	1.402	-2.6	82	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	0.117	0.122	-4.3	78	0.00
66 c	n-Nitrosodiphenylamine	0.675	0.668	1.0	81	0.00
67	4-Bromophenyl-phenylether	0.234	0.229	2.1	80	0.00
68	Hexachlorobenzene	0.254	0.260	-2.4	84	0.00
69	Atrazine	0.207	0.201	2.9	80	0.00
70 C	Pentachlorophenol	0.065	0.077	-18.5	88	0.00
71	Phenanthrene	1.117	1.116	0.1	83	0.00
72	Anthracene	1.111	1.120	-0.8	83	0.00
73	Carbazole	1.008	1.012	-0.4	84	0.00
74	Di-n-butylphthalate	1.298	1.274	1.8	82	0.00
75 C	Fluoranthene	1.143	1.167	-2.1	86	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
77	Benzydine	0.529	0.560	-5.9	96	0.00
78	Pyrene	1.738	1.784	-2.6	86	0.00
79 S	Terphenyl-d14	1.202	1.244	-3.5	89	0.00
80	Butylbenzylphthalate	0.735	0.723	1.6	85	0.00
81	Benzo(a)anthracene	1.376	1.364	0.9	89	0.00
82	3,3'-Dichlorobenzidine	0.426	0.411	3.5	89	0.00
83	Chrysene	1.288	1.293	-0.4	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.873	0.813	6.9	84	0.00
85 c	Di-n-octyl phthalate	1.204	1.098	8.8	83	0.00
86 I	Perylene-d12	1.000	1.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	1.373	1.541	-12.2	91	0.00
88	Benzo(b)fluoranthene	1.370	1.281	6.5	82	0.00
89	Benzo(k)fluoranthene	1.313	1.363	-3.8	98	0.00
90 C	Benzo(a)pyrene	1.078	1.091	-1.2	90	0.00
91	Dibenzo(a,h)anthracene	1.133	1.259	-11.1	89	0.00
92	Benzo(g,h,i)perylene	1.130	1.302	-15.2	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082922\
Data File : BF130032.D
Acq On : 29 Aug 2022 13:37
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 30 01:25:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF081422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 30 01:22:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0