

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082924\
 Data File : BF139223.D
 Acq On : 29 Aug 2024 16:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF082924

Quant Time: Aug 29 16:45:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 15:46:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	39.096	2.3	97	0.00
3	Pyridine	40.000	40.222	-0.6	100	0.00
4	n-Nitrosodimethylamine	40.000	39.672	0.8	98	0.00
5 S	2-Fluorophenol	80.000	79.660	0.4	99	0.00
6	Aniline	40.000	39.931	0.2	99	0.00
7 S	Phenol-d6	80.000	79.650	0.4	100	0.00
8	2-Chlorophenol	40.000	39.644	0.9	99	0.00
9	Benzaldehyde	40.000	40.428	-1.1	104	0.00
10 C	Phenol	40.000	40.347	-0.9	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.839	0.4	99	0.00
12	1,3-Dichlorobenzene	40.000	39.795	0.5	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.721	0.7	100	0.00
14	1,2-Dichlorobenzene	40.000	39.195	2.0	98	0.00
15	Benzyl Alcohol	40.000	40.203	-0.5	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.681	0.8	99	0.00
17	2-Methylphenol	40.000	39.976	0.1	98	0.00
18	Hexachloroethane	40.000	40.461	-1.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.272	1.8	100	0.00
20	3+4-Methylphenols	40.000	39.477	1.3	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	40.192	-0.5	100	0.00
23 S	Nitrobenzene-d5	80.000	83.440	-4.3	101	0.00
24	Nitrobenzene	40.000	41.186	-3.0	100	0.00
25	Isophorone	40.000	40.717	-1.8	101	0.00
26 C	2-Nitrophenol	40.000	44.722	-11.8	102	0.00
27	2,4-Dimethylphenol	40.000	40.504	-1.3	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.264	-0.7	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.926	-2.3	100	0.00
30	1,2,4-Trichlorobenzene	40.000	39.829	0.4	98	0.00
31	Naphthalene	40.000	39.762	0.6	98	0.00
32	Benzoic acid	40.000	44.894	-12.2	106	0.00
33	4-Chloroaniline	40.000	40.399	-1.0	100	0.00
34 C	Hexachlorobutadiene	40.000	40.403	-1.0	99	0.00
35	Caprolactam	40.000	41.349	-3.4	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.254	-3.1	101	0.00
37	2-Methylnaphthalene	40.000	40.308	-0.8	99	0.00
38	1-Methylnaphthalene	40.000	39.920	0.2	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.307	1.7	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.565	-3.9	97	0.00
42 S	2,4,6-Tribromophenol	80.000	83.000	-3.8	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.789	-2.0	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.576	-1.4	100	0.00
45 S	2-Fluorobiphenyl	80.000	78.588	1.8	100	0.00
46	1,1'-Biphenyl	40.000	40.052	-0.1	100	0.00
47	2-Chloronaphthalene	40.000	39.878	0.3	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082924\
 Data File : BF139223.D
 Acq On : 29 Aug 2024 16:07
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF082924

Quant Time: Aug 29 16:45:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 15:46:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.335	-8.3	102	0.00
49	Acenaphthylene	40.000	40.316	-0.8	101	0.00
50	Dimethylphthalate	40.000	40.629	-1.6	102	0.00
51	2,6-Dinitrotoluene	40.000	42.993	-7.5	103	0.00
52 C	Acenaphthene	40.000	40.569	-1.4	101	0.00
53	3-Nitroaniline	40.000	42.792	-7.0	104	0.00
54 P	2,4-Dinitrophenol	40.000	41.873	-4.7	113	0.00
55	Dibenzofuran	40.000	40.160	-0.4	102	0.00
56 P	4-Nitrophenol	40.000	44.147	-10.4	104	0.00
57	2,4-Dinitrotoluene	40.000	44.718	-11.8	106	0.00
58	Fluorene	40.000	39.783	0.5	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.488	-3.7	99	0.00
60	Diethylphthalate	40.000	40.559	-1.4	102	0.00
61	4-Chlorophenyl-phenylether	40.000	39.573	1.1	100	0.00
62	4-Nitroaniline	40.000	42.948	-7.4	103	0.00
63	Azobenzene	40.000	40.346	-0.9	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.985	-15.0	111	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.740	0.6	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.185	-0.5	100	0.00
68	Hexachlorobenzene	40.000	39.502	1.2	99	0.00
69	Atrazine	40.000	39.280	1.8	104	0.00
70 C	Pentachlorophenol	40.000	42.837	-7.1	101	0.00
71	Phenanthrene	40.000	39.796	0.5	102	0.00
72	Anthracene	40.000	39.874	0.3	101	0.00
73	Carbazole	40.000	41.139	-2.8	104	0.00
74	Di-n-butylphthalate	40.000	42.798	-7.0	109	0.00
75 C	Fluoranthene	40.000	41.817	-4.5	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	116	0.00
77	Benzydine	40.000	47.714	-19.3	119	0.00
78	Pyrene	40.000	39.717	0.7	108	0.00
79 S	Terphenyl-d14	80.000	78.847	1.4	108	0.00
80	Butylbenzylphthalate	40.000	46.435	-16.1	127	0.00
81	Benzo(a)anthracene	40.000	40.898	-2.2	114	0.00
82	3,3'-Dichlorobenzidine	40.000	40.347	-0.9	113	0.00
83	Chrysene	40.000	39.197	2.0	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.954	-4.9	116	0.00
85 c	Di-n-octyl phthalate	40.000	35.232	11.9	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.974	-2.4	99	0.00
88	Benzo(b)fluoranthene	40.000	41.807	-4.5	105	0.00
89	Benzo(k)fluoranthene	40.000	40.719	-1.8	105	0.00
90 C	Benzo(a)pyrene	40.000	41.265	-3.2	102	0.00
91	Dibenzo(a,h)anthracene	40.000	40.747	-1.9	98	0.00
92	Benzo(g,h,i)perylene	40.000	41.216	-3.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082924\
Data File : BF139223.D
Acq On : 29 Aug 2024 16:07
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF082924

Quant Time: Aug 29 16:45:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF082924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 15:46:29 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0