



Data Path : Z:\HPCHEM1\BNA F\DATA\BF083017\
Data File : BF098158.D
Acq On : 30 Aug 2017 16:16
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 09:16:03 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration



TIC: BF098158.D

(76) Benzidine



12.568min (-0.030) 0.00ng

response 320637

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.07#
183.00	12.20	11.09
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF083017\
Data File : BF098158.D
Acq On : 30 Aug 2017 16:16
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 09:16:03 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration



TIC: BF098158.D

(69) Pentachlorophenol (C)

11.045min (-0.024) 53784.58ng

response 37839

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	66.61
264.00	64.60	63.56
0.00	0.00	0.00