

Data Path : Z:\HPCHEM1\BNA F\Data\BF083117\
 Data File : BF098199.D
 Acq On : 31 Aug 2017 17:18
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 01:49:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 15:59:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.733	0.720	1.8	97	0.01
3	Pyridine	2.027	2.000	1.3	96	0.00
4	n-Nitrosodimethylamine	0.780	0.709	9.1	86	0.02
5 S	2-Fluorophenol	1.315	1.305	0.8	98	0.00
6	Aniline	2.510	2.417	3.7	95	0.00
7 S	Phenol-d6	1.787	1.813	-1.5	100	0.00
8	2-Chlorophenol	1.387	1.417	-2.2	99	0.00
9	Benzaldehyde	1.132	1.028	9.2	87	0.00
10 C	Phenol	2.089	2.084	0.2	94	0.00
11	bis(2-Chloroethyl)ether	1.577	1.573	0.3	98	0.00
12	1,3-Dichlorobenzene	1.492	1.498	-0.4	99	0.00
13 C	1,4-Dichlorobenzene	1.517	1.512	0.3	98	0.00
14	1,2-Dichlorobenzene	1.399	1.396	0.2	98	0.00
15	Benzyl Alcohol	1.338	1.257	6.1	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.122	1.997	5.9	92	0.00
17	2-Methylphenol	1.222	1.212	0.8	96	0.00
18	Hexachloroethane	0.595	0.600	-0.8	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.223	1.142	6.6	91	0.00
20	3+4-Methylphenols	1.493	1.444	3.3	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.528	0.494	6.4	93	0.00
23 S	Nitrobenzene-d5	0.368	0.406	-10.3	104	0.00
24	Nitrobenzene	0.410	0.438	-6.8	97	0.00
25	Isophorone	0.766	0.741	3.3	94	0.00
26 C	2-Nitrophenol	0.137	0.182	-32.8#	124	0.00
27	2,4-Dimethylphenol	0.319	0.320	-0.3	99	0.00
28	bis(2-Chloroethoxy)methane	0.503	0.499	0.8	96	0.00
29 C	2,4-Dichlorophenol	0.265	0.273	-3.0	98	0.00
30	1,2,4-Trichlorobenzene	0.294	0.292	0.7	97	0.00
31	Naphthalene	1.038	1.023	1.4	97	0.00
32	Benzoic acid	0.212	0.181	14.6	84	0.00
33	4-Chloroaniline	0.438	0.430	1.8	97	0.00
34 C	Hexachlorobutadiene	0.172	0.172	0.0	98	0.00
35	Caprolactam	0.090	0.092	-2.2	96	0.00
36 C	4-Chloro-3-methylphenol	0.341	0.337	1.2	94	0.00
37	2-Methylnaphthalene	0.637	0.627	1.6	96	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
39	1,2,4,5-Tetrachlorobenzene	0.614	0.611	0.5	97	0.00
40 P	Hexachlorocyclopentadiene	0.295	0.286	3.1	88	0.00
41 S	2,4,6-Tribromophenol	0.174	0.186	-6.9	100	0.00
42 C	2,4,6-Trichlorophenol	0.412	0.423	-2.7	97	0.00
43	2,4,5-Trichlorophenol	0.409	0.426	-4.2	98	0.00
44 S	2-Fluorobiphenyl	1.361	1.301	4.4	96	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF083117\
 Data File : BF098199.D
 Acq On : 31 Aug 2017 17:18
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 01:49:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 21 15:59:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.824	1.797	1.5	96	0.00
46	2-Chloronaphthalene	1.348	1.307	3.0	96	0.00
47	2-Nitroaniline	0.452	0.513	-13.5	104	0.00
48	Acenaphthylene	2.116	2.060	2.6	94	0.00
49	Dimethylphthalate	1.462	1.453	0.6	96	0.00
50	2,6-Dinitrotoluene	0.292	0.314	-7.5	100	0.00
51 C	Acenaphthene	1.335	1.227	8.1	90	0.00
52	3-Nitroaniline	0.376	0.412	-9.6	103	0.00
53 P	2,4-Dinitrophenol	0.106	0.145	-36.8#	129	0.00
54	Dibenzofuran	1.789	1.709	4.5	94	0.00
55 P	4-Nitrophenol	0.300	0.279	7.0	86	0.00
56	2,4-Dinitrotoluene	0.366	0.402	-9.8	100	0.00
57	Fluorene	1.383	1.348	2.5	97	0.00
58	2,3,4,6-Tetrachlorophenol	0.317	0.314	0.9	94	0.00
59	Diethylphthalate	1.529	1.431	6.4	90	0.00
60	4-Chlorophenyl-phenylether	0.642	0.633	1.4	97	0.00
61	4-Nitroaniline	0.358	0.406	-13.4	106	0.00
62	Azobenzene	1.816	1.651	9.1	88	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
64	4,6-Dinitro-2-methylphenol	0.083	0.113	-36.1#	126	0.00
65 c	n-Nitrosodiphenylamine	0.681	0.666	2.2	94	0.00
66	4-Bromophenyl-phenylether	0.208	0.210	-1.0	95	0.00
67	Hexachlorobenzene	0.212	0.214	-0.9	97	0.00
68	Atrazine	0.181	0.166	8.3	88	0.00
69 C	Pentachlorophenol	0.123	0.109	11.4	79	0.00
70	Phenanthrene	1.066	1.018	4.5	92	0.00
71	Anthracene	1.081	1.054	2.5	94	0.00
72	Carbazole	1.026	0.989	3.6	94	0.00
73	Di-n-butylphthalate	1.182	1.191	-0.8	94	0.00
74 C	Fluoranthene	1.112	1.062	4.5	92	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
76	Benzidine	0.656	0.545	16.9	83	0.00
77	Pyrene	1.636	1.607	1.8	93	0.00
78 S	Terphenyl-d14	0.959	0.946	1.4	95	0.00
79	Butylbenzylphthalate	0.627	0.680	-8.5	99	0.00
80	Benzo(a)anthracene	1.199	1.123	6.3	89	0.00
81	3,3'-Dichlorobenzidine	0.404	0.402	0.5	89	0.00
82	Chrysene	1.192	1.108	7.0	89	0.00
83	Bis(2-ethylhexyl)phthalate	0.695	0.827	-19.0	104	0.00
84 c	Di-n-octyl phthalate	1.209	1.438	-18.9	106	0.00
85	Indeno(1,2,3-cd)pyrene	0.877	0.956	-9.0	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Benzo(b)fluoranthene	1.261	1.194	5.3	90	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF083117\
Data File : BF098199.D
Acq On : 31 Aug 2017 17:18
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 01 01:49:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 21 15:59:04 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.184	1.193	-0.8	94	0.00
89 C	Benzo(a)pyrene	1.116	1.131	-1.3	95	0.00
90	Dibenzo(a,h)anthracene	0.909	1.037	-14.1	106	0.00
91	Benzo(a,h,i)perylene	0.948	1.075	-13.4	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1