



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083121\
 Data File : BF125279.D
 Acq On : 31 Aug 2021 8:40
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 24 12:39:21 2021






AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	66899	PASS
68	69	0.00	2	1.7	935	PASS
69	198	0.00	100	33.8	54224	PASS
70	69	0.00	2	0.4	216	PASS
127	198	10	80	42.9	68752	PASS
197	198	0.00	2	0.2	369	PASS
198	198	100	100	100.0	160384	PASS
199	198	5	9	6.7	10822	PASS
275	198	10	60	27.1	43421	PASS
365	198	1	100	3.6	5799	PASS
441	198	0.01	100	16.0	25617	PASS
442	442	50	100	100.0	168696	PASS
443	442	15	24	19.1	32199	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083121\
 Data File : BF125279.D
 Acq On : 31 Aug 2021 8:40
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 31 10:36:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration



TIC: BF125279.D\data.ms

(70) Pentachlorophenol (C)

11.245min (+ 0.006) 24214.00 ng




response 128198

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	70.97
264.00	69.90	65.67
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF083121\
 Data File : BF125279.D
 Acq On : 31 Aug 2021 8:40
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 31 10:36:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration



TIC: BF125279.D\data.ms

(77) Benzidine

12.780min (+ 0.006) 170918.56 ng

response 728254

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.23
183.00	10.10	11.24
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/31/2021	BNA_F	<u>BF125279.D</u>
Compound Name	Response	Retention Time
DDT	327346	13.59
DDD	5197	13.29
DDE	1133	12.96
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6330	333676	1.90

Instrument :
BNA_F
ClientSampleId :
DFTPP