

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090115\
 Data File : BF081317.D
 Acq On : 1 Sep 2015 22:09
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 02 13:31:14 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 01 17:35:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.579	0.574	0.9	102	0.00
3	Pyridine	1.596	1.574	1.4	87	0.00
4	n-Nitrosodimethylamine	0.887	0.869	2.0	94	0.00
5 S	2-Fluorophenol	1.289	1.277	0.9	106	0.00
6	Aniline	2.410	2.269	5.9	96	0.00
7 S	Phenol-d6	1.706	1.747	-2.4	103	0.00
8	2-Chlorophenol	1.420	1.438	-1.3	103	0.00
9	Benzaldehyde	1.069	1.040	2.7	99	0.00
10 C	Phenol	1.979	2.152	-8.7	100	0.00
11	bis(2-Chloroethyl)ether	1.428	1.368	4.2	97	-0.01
12	1,3-Dichlorobenzene	1.518	1.463	3.6	100	0.00
13 C	1,4-Dichlorobenzene	1.545	1.509	2.3	100	0.00
14	1,2-Dichlorobenzene	1.391	1.374	1.2	103	0.00
15	Benzyl Alcohol	1.400	1.409	-0.6	97	0.00
16	2,2'-oxybis(1-Chloropropane	2.736	2.641	3.5	103	0.00
17	2-Methylphenol	1.133	1.200	-5.9	106	0.00
18	Hexachloroethane	0.601	0.611	-1.7	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.161	1.104	4.9	101	0.00
20	3+4-Methylphenols	1.528	1.527	0.1	103	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.546	0.561	-2.7	101	0.00
23 S	Nitrobenzene-d5	0.415	0.422	-1.7	101	0.00
24	Nitrobenzene	0.430	0.433	-0.7	97	0.00
25	Isophorone	0.771	0.772	-0.1	103	0.00
26 C	2-Nitrophenol	0.188	0.170	9.6	96	0.00
27	2,4-Dimethylphenol	0.324	0.353	-9.0	99	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.499	-12.1	101	0.00
29 C	2,4-Dichlorophenol	0.281	0.296	-5.3	103	0.00
30	1,2,4-Trichlorobenzene	0.305	0.310	-1.6	98	0.00
31	Naphthalene	1.067	1.115	-4.5	106	0.00
32	Benzoic acid	0.221	0.227	-2.7	94	0.00
33	4-Chloroaniline	0.435	0.451	-3.7	92	0.00
34 C	Hexachlorobutadiene	0.186	0.187	-0.5	106	0.00
35	Caprolactam	0.086	0.082	4.7	92	0.00
36 C	4-Chloro-3-methylphenol	0.315	0.311	1.3	100	0.00
37	2-Methylnaphthalene	0.694	0.641	7.6	100	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
39	1,2,4,5-Tetrachlorobenzene	0.633	0.684	-8.1	103	0.00
40 P	Hexachlorocyclopentadiene	0.310	0.326	-5.2	100	0.00
41 S	2,4,6-Tribromophenol	0.178	0.193	-8.4	101	0.00
42 C	2,4,6-Trichlorophenol	0.437	0.435	0.5	99	-0.01
43	2,4,5-Trichlorophenol	0.445	0.486	-9.2	104	0.00
44 S	2-Fluorobiphenyl	1.561	1.700	-8.9	102	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090115\
 Data File : BF081317.D
 Acq On : 1 Sep 2015 22:09
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 02 13:31:14 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 01 17:35:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.840	1.724	6.3	102	0.00
46	2-Chloronaphthalene	1.329	1.393	-4.8	96	0.00
47	2-Nitroaniline	0.542	0.582	-7.4	101	0.00
48	Acenaphthylene	2.357	2.172	7.8	98	0.00
49	Dimethylphthalate	1.554	1.640	-5.5	107	0.00
50	2,6-Dinitrotoluene	0.353	0.392	-11.0	102	0.00
51 C	Acenaphthene	1.341	1.216	9.3	99	0.00
52	3-Nitroaniline	0.402	0.427	-6.2	102	0.00
53 P	2,4-Dinitrophenol	0.149	0.161	-8.1	99	0.00
54	Dibenzofuran	1.958	1.868	4.6	101	0.00
55 P	4-Nitrophenol	0.252	0.294	-16.7	97	0.00
56	2,4-Dinitrotoluene	0.449	0.462	-2.9	100	0.00
57	Fluorene	1.486	1.691	-13.8	104	0.00
58	2,3,4,6-Tetrachlorophenol	0.340	0.354	-4.1	105	0.00
59	Diethylphthalate	1.635	1.732	-5.9	103	0.00
60	4-Chlorophenyl-phenylether	0.693	0.709	-2.3	104	0.00
61	4-Nitroaniline	0.406	0.391	3.7	100	0.00
62	Azobenzene	1.817	1.994	-9.7	101	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
64	4,6-Dinitro-2-methylphenol	0.112	0.131	-17.0	104	0.00
65 c	n-Nitrosodiphenylamine	0.728	0.722	0.8	102	0.00
66	4-Bromophenyl-phenylether	0.202	0.184	8.9	105	0.00
67	Hexachlorobenzene	0.211	0.196	7.1	102	0.00
68	Atrazine	0.211	0.215	-1.9	108	0.00
69 C	Pentachlorophenol	0.082	0.086	-4.9	107	0.00
70	Phenanthrene	1.112	1.108	0.4	99	0.00
71	Anthracene	1.142	1.089	4.6	101	0.00
72	Carbazole	1.072	0.953	11.1	98	0.00
73	Di-n-butylphthalate	1.266	1.187	6.2	102	0.00
74 C	Fluoranthene	1.204	1.108	8.0	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
76	Benzidine	0.859	0.761	11.4	103	0.00
77	Pyrene	1.481	1.462	1.3	103	0.00
78 S	Terphenyl-d14	0.859	0.757	11.9	109	0.00
79	Butylbenzylphthalate	0.644	0.596	7.5	103	0.00
80	Benzo(a)anthracene	1.274	1.240	2.7	102	0.00
81	3,3'-Dichlorobenzidine	0.430	0.364	15.3	99	0.00
82	Chrysene	1.142	0.881	22.9	100	0.00
83	Bis(2-ethylhexyl)phthalate	0.850	0.768	9.6	106	0.00
84 c	Di-n-octyl phthalate	1.400	1.252	10.6	102	0.00
85	Indeno(1,2,3-cd)pyrene	0.838	0.757	9.7	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Benzo(b)fluoranthene	1.428	1.704	-19.3	98	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090115\
Data File : BF081317.D
Acq On : 1 Sep 2015 22:09
Operator : UM/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 02 13:31:14 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 01 17:35:35 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.185	1.029	13.2	106	0.00
89 C	Benzo(a)pyrene	1.210	1.185	2.1	104	0.00
90	Dibenzo(a,h)anthracene	0.935	0.909	2.8	98	0.00
91	Benzo(g,h,i)perylene	0.892	0.892	0.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0