

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090117\  
 Data File : BF098216.D  
 Acq On : 1 Sep 2017 10:29  
 Operator : SJ/JU  
 Sample : PB101969BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB101969BL

Quant Time: Sep 01 13:04:01 2017  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 01 12:59:02 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.72  | 152  | 108788   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.00  | 136  | 439311   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.75  | 164  | 206642   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 11.23 | 188  | 429459   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.87 | 240  | 325016   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.28 | 264  | 232089   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.34  | 112  | 897972   | 125.55 | ng    | 0.01     |
| 7) Phenol-d6                | 6.37  | 99   | 1221327  | 125.68 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.29  | 82   | 747555   | 92.44  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.55 | 330  | 258220   | 144.02 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.08  | 172  | 1163058  | 82.69  | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.82 | 244  | 1115050  | 71.54  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090117\  
Data File : BF098216.D  
Acq On : 1 Sep 2017 10:29  
Operator : SJ/JU  
Sample : PB101969BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB101969BL

Quant Time: Sep 01 13:04:01 2017  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 01 12:59:02 2017  
Response via : Initial Calibration

