

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090117\
 Data File : BF098224.D
 Acq On : 1 Sep 2017 14:08
 Operator : SJ/JU
 Sample : I5006-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-14-COMP

Manual Integrations
 APPROVED

mohammad
 9/5/2017 5:22:04 PM

Quant Time: Sep 01 16:40:18 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 01 16:30:12 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	113762	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	460067	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	196497	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	316480	20.00	ng	0.00
75) Chrysene-d12	13.87	240	226251	20.00	ng	0.00
86) Perylene-d12	15.29	264	205703	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	920121	123.03	ng	0.00
7) Phenol-d6	6.37	99	1258466	123.84	ng	0.00
23) Nitrobenzene-d5	7.29	82	780872	92.21	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	208039	122.02	ng	0.00
44) 2-Fluorobiphenyl	9.08	172	1047087	78.29	ng	0.00
78) Terphenyl-d14	12.82	244	677773	62.47	ng	0.00
Target Compounds						
10) Phenol	6.38	94	47706	4.014	ng	87
49) Dimethylphthalate	9.47	163	199467	13.887	ng	99
57) Fluorene	10.30	166	27880	2.052	ng	# 97
70) Phenanthrene	11.26	178	308092	18.269	ng	99
71) Anthracene	11.31	178	85027	4.971	ng	98
74) Fluoranthene	12.45	202	342825	19.476	ng	99
77) Pyrene	12.67	202	308698	16.682	ng	98
80) Benzo(a)anthracene	13.86	228	159486	11.761	ng	98
82) Chrysene	13.89	228	131718	9.765	ng	98
87) Benzo(b)fluoranthene	14.88	252	155219m	11.971	ng	
88) Benzo(k)fluoranthene	14.90	252	55765m	4.578	ng	
89) Benzo(a)pyrene	15.23	252	112328	9.786	ng	98
91) Benzo(a,h,i)perylene	17.06	276	45176	4.633	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090117\
Data File : BF098224.D
Acq On : 1 Sep 2017 14:08
Operator : SJ/JU
Sample : I5006-04
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-14-COMP

Manual Integrations
APPROVED

mohammad
9/5/2017 5:22:04 PM

Quant Time: Sep 01 16:40:18 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 01 16:30:12 2017
Response via : Initial Calibration

