


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF090117\
 Data File : BF098213.D
 Acq On : 1 Sep 2017 8:30
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 21 15:59:04 2017



AutoFind: Scans 1599, 1600, 1601; Background Corrected with Scan 1594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	31828	PASS
68	69	0.00	2	1.9	661	PASS
69	198	0.00	100	42.2	34722	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.0	43621	PASS
197	198	0.00	2	0.6	496	PASS
198	198	100	100	100.0	82253	PASS
199	198	5	9	6.6	5438	PASS
275	198	10	60	23.7	19520	PASS
365	198	1	100	3.0	2465	PASS
441	442	0.01	24	15.9	12189	PASS
442	198	50	100	93.2	76672	PASS
443	442	15	24	18.7	14341	PASS