

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090121\
 Data File : BF125305.D
 Acq On : 1 Sep 2021 14:28
 Operator : JU/CG
 Sample : M3561-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OUII-C

Quant Time: Sep 02 08:15:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 12:39:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	168437	20.000	ng	-0.01
21) Naphthalene-d8	8.186	136	663989	20.000	ng	#-0.01
39) Acenaphthene-d10	9.945	164	367383	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	706544	20.000	ng	0.00
76) Chrysene-d12	14.074	240	588989	20.000	ng	# 0.00
86) Perylene-d12	15.574	264	470012	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	639286	57.447	ng	0.00
7) Phenol-d6	6.528	99	522182	36.282	ng	-0.01
23) Nitrobenzene-d5	7.475	82	1114174	84.477	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	572355	156.070	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	1923768	72.981	ng	0.00
79) Terphenyl-d14	13.027	244	2562383	69.307	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090121\
Data File : BF125305.D
Acq On : 1 Sep 2021 14:28
Operator : JU/CG
Sample : M3561-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OUII-C

Quant Time: Sep 02 08:15:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 24 12:39:21 2021
Response via : Initial Calibration

