

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090215\
 Data File : BF081370.D
 Acq On : 3 Sep 2015 2:25
 Operator : UM/IZ
 Sample : G3490-01 5X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 28TP22(4.5-5)

Manual Integrations
 APPROVED

MMDadoda
 9/3/2015 1:45:27 PM

Quant Time: Sep 03 03:41:07 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 01 17:35:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.39	152	56139	20.00	ng	0.00
21) Naphthalene-d8	7.68	136	224051	20.00	ng	0.00
38) Acenaphthene-d10	9.42	164	98764	20.00	ng	0.00
63) Phenanthrene-d10	10.88	188	176412	20.00	ng	0.00
75) Chrysene-d12	13.50	240	113867	20.00	ng	0.00
86) Perylene-d12	14.81	264	85709	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.94	112	91901	25.40	ng	0.01
7) Phenol-d6	6.06	99	133310	27.83	ng	0.00
23) Nitrobenzene-d5	6.96	82	79893	17.20	ng	-0.01
41) 2,4,6-Tribromophenol	10.19	330	18602	21.15	ng	-0.01
44) 2-Fluorobiphenyl	8.75	172	113785	14.76	ng	-0.01
78) Terphenyl-d14	12.47	244	84455	17.27	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.16	163	17336	2.26	ng	# 96
70) Phenanthrene	10.90	178	44641	4.55	ng	97
74) Fluoranthene	12.08	202	136357	12.84	ng	88
77) Pyrene	12.31	202	122972	14.58	ng	99
79) Butylbenzylphthalate	12.96	149	19110	5.21	ng	# 81
80) Benzo(a)anthracene	13.49	228	40556	5.59	ng	# 92
82) Chrysene	13.52	228	24584m	3.78	ng	
83) Bis(2-ethylhexyl)phthalate	13.52	149	111671	23.07	ng	# 90
87) Benzo(b)fluoranthene	14.48	252	23294m	3.81	ng	
89) Benzo(a)pyrene	14.75	252	16792	3.24	ng	# 75
91) Benzo(g,h,i)perylene	16.22	276	7755	2.03	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090215\
Data File : BF081370.D
Acq On : 3 Sep 2015 2:25
Operator : UM/IZ
Sample : G3490-01 5X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
28TP22(4.5-5)

Manual Integrations
APPROVED

MMDadoda
9/3/2015 1:45:27 PM

Quant Time: Sep 03 03:41:07 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 01 17:35:35 2015
Response via : Initial Calibration

