


Data Path : Z:\HPCHEM1\BNA_F\Data\BF090217\
 Data File : BF098243.D
 Acq On : 2 Sep 2017 13:57
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 05 17:28:20 2017



AutoFind: Scans 1597, 1598, 1599; Background Corrected with Scan 1593

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	35827	PASS
68	69	0.00	2	1.7	681	PASS
69	198	0.00	100	44.4	40589	PASS
70	69	0.00	2	0.3	140	PASS
127	198	10	80	52.7	48216	PASS
197	198	0.00	2	1.0	950	PASS
198	198	100	100	100.0	91514	PASS
199	198	5	9	6.6	6046	PASS
275	198	10	60	23.4	21382	PASS
365	198	1	100	2.8	2536	PASS
441	442	0.01	24	15.9	13130	PASS
442	198	50	100	90.3	82661	PASS
443	442	15	24	19.8	16330	PASS