

Data Path : Z:\HPCHEM1\BNA F\DATA\BF090315\
 Data File : BF081395.D
 Acq On : 4 Sep 2015 00:58
 Operator : UM/IZ
 Sample : G3511-12MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Quant Time: Sep 04 03:07:03 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 02:03:29 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.36	152	57147	20.00	ng	-0.01
21) Naphthalene-d8	7.66	136	216968	20.00	ng	-0.01
38) Acenaphthene-d10	9.40	164	104671	20.00	ng	0.00
63) Phenanthrene-d10	10.87	188	203135	20.00	ng	0.00
75) Chrysene-d12	13.47	240	163985	20.00	ng	0.00
86) Perylene-d12	14.78	264	107495	20.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	4.92	112	499917	135.73	ng	0.01
7) Phenol-d6	6.04	99	706225	144.85	ng	0.00
23) Nitrobenzene-d5	6.95	82	424361	94.36	ng	0.00
41) 2,4,6-Tribromophenol	10.19	330	147700	158.48	ng	0.00
44) 2-Fluorobiphenyl	8.74	172	646115	79.10	ng	-0.01
78) Terphenyl-d14	12.45	244	654554	92.92	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	1.76	88	61724	37.31	ng	# 88
3) Pyridine	2.30	79	164906	36.15	ng	# 94
4) n-Nitrosodimethylamine	2.26	42	108797	42.93	ng	# 76
6) Aniline	6.03	93	119218	17.32	ng	# 70
8) 2-Chlorophenol	6.16	128	191590	47.20	ng	# 89
9) Benzaldehyde	5.91	77	23604	7.73	ng	# 81
10) Phenol	6.06	94	210521	37.23	ng	# 45
11) bis(2-Chloroethyl)ether	6.12	93	179991	44.12	ng	# 88
12) 1,3-Dichlorobenzene	6.31	146	186780	43.07	ng	# 91
13) 1,4-Dichlorobenzene	6.39	146	191149	43.29	ng	# 91
14) 1,2-Dichlorobenzene	6.54	146	159711	40.17	ng	# 87
15) Benzyl Alcohol	6.54	79	168742	42.19	ng	# 71
16) 2,2'-oxybis(1-Chloropropan	6.67	45	354994	45.40	ng	# 9
17) 2-Methylphenol	6.67	107	154532	47.72	ng	# 76
18) Hexachloroethane	6.88	117	73625	42.86	ng	# 90
19) n-Nitroso-di-n-propylamine	6.81	70	145658	43.90	ng	# 88
20) 3+4-Methylphenols	6.83	107	200224	45.85	ng	# 89
22) Acetophenone	6.80	105	272687	46.01	ng	# 78
24) Nitrobenzene	6.97	77	237949	50.95	ng	# 66
25) Isophorone	7.21	82	369150	44.13	ng	# 82
26) 2-Nitrophenol	7.29	139	99573	48.92	ng	# 59
27) 2,4-Dimethylphenol	7.35	122	144904	41.22	ng	# 74
28) bis(2-Chloroethoxy)methane	7.44	93	216891	44.89	ng	# 92
29) 2,4-Dichlorophenol	7.54	162	147144	48.27	ng	# 97
30) 1,2,4-Trichlorobenzene	7.61	180	157525	47.56	ng	# 98
31) Naphthalene	7.68	128	484452	41.87	ng	# 98
32) Benzoic acid	7.45	122	4902	4.39	ng	# 73
33) 4-Chloroaniline	7.75	127	55543	11.77	ng	# 86
34) Hexachlorobutadiene	7.80	225	87501	43.41	ng	# 96
35) Caprolactam	8.14	113	24668	26.38	ng	# 35
36) 4-Chloro-3-methylphenol	8.26	107	174449	51.12	ng	# 74
37) 2-Methylnaphthalene	8.38	142	354443	47.07	ng	# 91
39) 1,2,4,5-Tetrachlorobenzene	8.55	216	136454	41.22	ng	# 99
40) Hexachlorocyclopentadiene	8.52	237	143205	88.15	ng	# 95

Data Path : Z:\HPCHEM1\BNA F\DATA\BF090315\
 Data File : BF081395.D
 Acq On : 4 Sep 2015 00:58
 Operator : UM/IZ
 Sample : G3511-12MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Quant Time: Sep 04 03:07:03 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 02:03:29 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.66	196	95727	41.90	nq	99
43) 2,4,5-Trichlorophenol	8.72	196	104311	44.76	nq #	96
45) 1,1'-Biphenyl	8.84	154	425094	44.14	nq	94
46) 2-Chloronaphthalene	8.86	162	291437	41.91	nq #	89
47) 2-Nitroaniline	8.97	65	142199	50.17	nq #	65
48) Acenaphthylene	9.27	152	544667	44.15	nq	98
49) Dimethylphthalate	9.16	163	542367	66.69	nq #	98
50) 2,6-Dinitrotoluene	9.22	165	80489	43.61	nq #	94
51) Acenaphthene	9.44	154	318212	45.34	nq	88
52) 3-Nitroaniline	9.38	138	55481	26.40	nq #	77
53) 2,4-Dinitrophenol	9.48	184	16561	21.30	nq #	1
54) Dibenzofuran	9.61	168	469387	45.80	nq #	89
55) 4-Nitrophenol	9.61	139	299505	226.88	nq #	31
56) 2,4-Dinitrotoluene	9.61	165	95509	40.64	nq #	7
57) Fluorene	9.94	166	326238	41.95	nq	95
58) 2,3,4,6-Tetrachlorophenol	9.74	232	86948	48.86	nq	92
59) Diethylphthalate	9.85	149	363188	42.46	nq	94
60) 4-Chlorophenyl-phenylether	9.95	204	171060	47.19	nq	96
61) 4-Nitroaniline	9.99	138	79124	37.27	nq #	26
62) Azobenzene	10.10	77	403231	42.40	nq #	81
64) 4,6-Dinitro-2-methylphenol	10.02	198	45195	39.78	nq	85
65) n-Nitrosodiphenylamine	10.08	169	316087	42.76	nq	91
66) 4-Bromophenyl-phenylether	10.43	248	94858	46.18	nq	98
67) Hexachlorobenzene	10.49	284	96695	45.03	nq #	80
68) Atrazine	10.62	200	106429	49.76	nq	82
69) Pentachlorophenol	10.68	266	88479	84.12	nq	99
70) Phenanthrene	10.89	178	448465	39.71	nq	97
71) Anthracene	10.95	178	516938	44.55	nq	99
72) Carbazole	11.11	167	470721	43.23	nq	96
73) Di-n-butylphthalate	11.46	149	572033	44.49	nq #	98
74) Fluoranthene	12.07	202	535785	43.83	nq	85
76) Benzidine	12.20	184	117421	16.68	nq	98
77) Pyrene	12.28	202	536937	44.20	nq	98
79) Butylbenzylphthalate	12.94	149	242961	45.99	nq #	81
80) Benzo(a)anthracene	13.46	228	432342	41.40	nq	97
81) 3,3'-Dichlorobenzidine	13.45	252	69917	19.85	nq #	92
82) Chrysene	13.51	228	385924	41.23	nq	96
83) Bis(2-ethylhexyl)phthalate	13.51	149	312100	44.77	nq #	94
84) Di-n-octyl phthalate	14.11	149	482624	42.04	nq	99
85) Indeno(1,2,3-cd)pyrene	15.86	276	264106	38.45	nq #	86
87) Benzo(b)fluoranthene	14.48	252	638823	83.24	nq #	88
88) Benzo(k)fluoranthene	14.48	252	638823	100.32	nq #	88
89) Benzo(a)pyrene	14.73	252	272923	41.97	nq #	88
90) Dibenzo(a,h)anthracene	15.87	278	225709	44.92	nq #	80
91) Benzo(g,h,i)perylene	16.18	276	220711	46.02	nq #	73


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF090315\
 Data File : BF081395.D
 Acq On : 4 Sep 2015 00:58
 Operator : UM/IZ
 Sample : G3511-12MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Quant Time: Sep 04 03:07:03 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 04 02:03:29 2015
 Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.36 min Scan# 418

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:152 Resp: 57147

Ion Ratio Lower Upper

152 100

150 162.1 124.7 187.1




115 70.9 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 37.31 ng

RT: 1.76 min Scan# 15

Delta R.T. 0.08 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion: 88 Resp: 61724

Ion Ratio Lower Upper

88 100

58 108.5 77.7 116.5



43 41.9 26.6 40.0#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0









#3
 Pvridine
 Concen: 36.15 ng
 RT: 2.30 min Scan# 62
 Delta R.T. 0.06 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion: 79 Resp: 164906
 Ion Ratio Lower Upper
 79 100
 52 100.7 76.8 115.2
 51 46.2 32.2 48.4



#4
 n-Nitrosodimethylamine
 Concen: 42.93 ng
 RT: 2.26 min Scan# 59
 Delta R.T. 0.07 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion: 42 Resp: 108797
 Ion Ratio Lower Upper
 42 100
 74 102.0 105.0 157.6#
 44 7.9 6.6 10.0





#5

2-Fluorophenol

Concen: 135.73 ng

RT: 4.92 min Scan# 292

Delta R.T. 0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:112 Resp: 499917

Ion Ratio Lower Upper

112 100

64 79.1 39.7 59.5#




63 42.0 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 17.32 ng

RT: 6.03 min Scan# 389

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion: 93 Resp: 119218

Ion Ratio Lower Upper

93 100

66 55.4 27.2 40.8#



65 25.4 14.7 22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 144.85 ng

RT: 6.04 min Scan# 390

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD





Tot Ion: 99 Resp: 706225

Ion Ratio Lower Upper

99 100

42 22.1 13.7 20.5#

71 37.8 14.2 21.2#



#8

2-Chlorophenol

Concen: 47.20 ng

RT: 6.16 min Scan# 400

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58




Tgt Ion:128 Resp: 191590


Ion Ratio Lower Upper

128 100

130 32.4 12.2 52.2

64 52.8 20.5 60.5









#9
Benzaldehyde
Concen: 7.73 ng
RT: 5.91 min Scan# 378
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion: 77 Resp: 23604
Ion Ratio Lower Upper
77 100
105 97.0 91.6 131.6
106 96.8 102.6 142.6#



#10
Phenol
Concen: 37.23 ng
RT: 6.06 min Scan# 391
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion: 94 Resp: 210521
Ion Ratio Lower Upper
94 100
65 37.5 0.7 40.7
66 55.1 0.8 40.8#









#11
bis(2-Chloroethyl)ether
Concen: 44.12 ng
RT: 6.12 min Scan# 397
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion: 93 Resp: 179991
Ion Ratio Lower Upper
93 100
63 100.7 65.6 105.6
95 33.1 13.6 53.6



#12
1,3-Dichlorobenzene
Concen: 43.07 ng
RT: 6.31 min Scan# 413
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion: 146 Resp: 186780
Ion Ratio Lower Upper
146 100
148 64.9 51.0 76.4
75 33.5 15.0 22.6#





#13

1,4-Dichlorobenzene

Concen: 43.29 ng

RT: 6.39 min Scan# 420

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:146 Resp: 191149

Ion Ratio Lower Upper

146 100

148 64.4 51.8 77.6




111 45.5 24.9 37.3#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 40.17 ng

RT: 6.54 min Scan# 433

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:146 Resp: 159711

Ion Ratio Lower Upper

146 100

148 63.7 52.7 79.1



111 54.4 28.8 43.2#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 42.19 ng

RT: 6.54 min Scan# 433

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion: 79 Resp: 168742

Ion Ratio Lower Upper

79 100

108 68.2 88.6 132.8#




77 64.7 47.0 70.4



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-chloropropane)

Concen: 45.40 ng

RT: 6.67 min Scan# 445

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion: 45 Resp: 354994

Ion Ratio Lower Upper

45 100

77 39.8 0.0 28.1#



79 37.1 0.0 26.0#




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0






#17
2-Methylphenol
Concen: 47.72 ng
RT: 6.67 min Scan# 445
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD

Tot Ion	Ratio	Lower	Upper
107	100		
108	115.8	96.6	145.0
77	65.2	23.3	34.9#
79	60.7	22.0	33.0#






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E


Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18
Hexachloroethane
Concen: 42.86 ng
RT: 6.88 min Scan# 463
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
117	100		
119	93.0	83.8	125.6
201	76.0	55.1	82.7








Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 43.90 ng
RT: 6.81 min Scan# 457
Delta R.T. -0.01 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58


Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion: 70 Resp: 145658
Ion Ratio Lower Upper
70 100
42 76.7 63.8 95.6
101 7.9 9.2 13.8#
130 15.7 27.3 40.9#



#20
3+4-Methylphenols
Concen: 45.85 ng
RT: 6.83 min Scan# 459
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:107 Resp: 200224
Ion Ratio Lower Upper
107 100
108 95.0 74.6 114.6
77 37.2 0.7 40.7
79 35.1 0.0 38.9





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.66 min Scan# 531

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD

Tot Ion:136 Resp: 216968


Ion Ratio Lower Upper

136 100

137 11.4 9.0 13.6

54 13.1 8.2 12.2#

68 6.0 5.8 8.6






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 46.01 ng

RT: 6.80 min Scan# 456

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Tgt Ion:105 Resp: 272687


Ion Ratio Lower Upper

105 100

71 1.8 4.7 7.1#

51 34.8 22.0 33.0#

120 20.4 30.1 45.1#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 94.36 ng
 RT: 6.95 min Scan# 469
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion: 82 Resp: 424361
 Ion Ratio Lower Upper
 82 100
 128 38.5 51.0 76.6#
 54 64.8 47.5 71.3



#24
 Nitrobenzene
 Concen: 50.95 ng
 RT: 6.97 min Scan# 471
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion: 77 Resp: 237949
 Ion Ratio Lower Upper
 77 100
 123 42.2 59.8 89.8#
 65 14.9 10.2 15.4





#25

Isophorone

Concen: 44.13 ng

RT: 7.21 min Scan# 492

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion: 82 Resp: 369150

Ion Ratio Lower Upper




82 100

95 7.0 4.0 6.0#

138 12.7 18.3 27.5#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 48.92 ng

RT: 7.29 min Scan# 499

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion: 139 Resp: 99573

Ion Ratio Lower Upper



139 100


109 31.2 11.0 16.4#

65 53.1 24.7 37.1#



Abundance Ion 139.00 (138.70 to 139.70): BF0
Ion 109.00 (108.70 to 109.70): BF0
Ion 65.00 (64.70 to 65.70): BF0









#27
 2,4-Dimethylphenol
 Concen: 41.22 ng
 RT: 7.35 min Scan# 504
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:122 Resp: 144904
 Ion Ratio Lower Upper
 122 100
 107 125.0 71.8 107.6#
 121 56.8 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 44.89 ng
 RT: 7.44 min Scan# 512
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion: 93 Resp: 216891
 Ion Ratio Lower Upper
 93 100
 95 32.6 27.5 41.3
 123 8.8 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 48.27 ng
 RT: 7.54 min Scan# 521
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Tot Ion	Ratio	Lower	Upper
162	100		
164	65.8	44.9	84.9
98	36.9	13.0	53.0




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 47.56 ng
 RT: 7.61 min Scan# 527
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.2	77.1	115.7
145	30.7	23.3	34.9



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 41.87 ng
RT: 7.68 min Scan# 533
Delta R.T. -0.01 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion:128 Resp: 484452
Ion Ratio Lower Upper
128 100
129 11.3 8.4 12.6
127 13.2 9.9 14.9



#32
Benzoic acid
Concen: 4.39 ng
RT: 7.45 min Scan# 513
Delta R.T. -0.06 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:122 Resp: 4902
Ion Ratio Lower Upper
122 100
105 113.1 76.1 116.1
77 92.4 40.5 80.5#









#33
4-Chloroaniline
Concen: 11.77 ng
RT: 7.75 min Scan# 539
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD

Tot Ion:	127	Resp:	55543
Ion	Ratio	Lower	Upper
127	100		
129	31.6	25.8	38.6
65	36.6	17.9	26.9#
92	20.9	10.5	15.7#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 43.41 ng
RT: 7.80 min Scan# 544
Delta R.T. -0.01 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:	225	Resp:	87501
Ion	Ratio	Lower	Upper
225	100		
223	65.8	50.2	75.2
227	61.3	52.2	78.2



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 26.38 ng

RT: 8.14 min Scan# 573

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:113 Resp: 24668

Ion Ratio Lower Upper

113 100

55 261.1 152.4 192.4#




56 199.1 104.6 144.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 51.12 ng

RT: 8.26 min Scan# 584

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:107 Resp: 174449

Ion Ratio Lower Upper

107 100

144 21.3 25.4 38.0#



142 69.4 77.3 115.9#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 47.07 ng
 RT: 8.38 min Scan# 594
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:142 Resp: 354443
 Ion Ratio Lower Upper
 142 100
 141 87.9 67.5 101.3
 115 37.3 18.2 27.2#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.40 min Scan# 684
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:164 Resp: 104671
 Ion Ratio Lower Upper
 164 100
 162 107.3 75.2 112.8
 160 45.9 31.5 47.3





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.22 ng

RT: 8.55 min Scan# 609

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD

Tot Ion:216 Resp: 136454



Ion Ratio Lower Upper

216 100

214 78.5 63.5 95.3

179 21.4 16.6 24.8

108 21.7 16.3 24.5





Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 88.15 ng

RT: 8.52 min Scan# 607

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58



Tgt Ion:237 Resp: 143205

Ion Ratio Lower Upper

237 100

235 63.7 42.0 82.0


272 10.2 0.0 36.1




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E





#41

2,4,6-Tribromophenol

Concen: 158.48 ng

RT: 10.19 min Scan# 753

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:330 Resp: 147700

Ion Ratio Lower Upper

330 100

332 95.6 76.6 114.8




141 40.1 29.7 44.5



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#42

2,4,6-Trichlorophenol

Concen: 41.90 ng

RT: 8.66 min Scan# 619

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:196 Resp: 95727

Ion Ratio Lower Upper

196 100

198 93.3 75.7 113.5



200 28.8 23.9 35.9




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 44.76 ng
 RT: 8.72 min Scan# 624
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Tot Ion:196	Resp:	104311
Ion Ratio	Lower	Upper
196	100	
198	93.7	75.4 113.0
97	38.4	33.3 49.9
132	23.9	24.6 37.0#




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 79.10 ng
 RT: 8.74 min Scan# 626
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:172	Resp:	646115
Ion Ratio	Lower	Upper
172	100	
171	36.0	26.1 39.1
170	23.8	17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 44.14 ng
 RT: 8.84 min Scan# 635
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Tot Ion:154 Resp: 425094
 Ion Ratio Lower Upper
 154 100
 153 40.5 17.1 57.1
 76 9.6 0.0 31.6




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 41.91 ng
 RT: 8.86 min Scan# 636
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:162 Resp: 291437
 Ion Ratio Lower Upper
 162 100
 127 45.9 27.6 41.4#
 164 32.2 25.4 38.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 50.17 ng
RT: 8.97 min Scan# 646
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion: 65 Resp: 142199
Ion Ratio Lower Upper
65 100
92 62.8 55.4 83.0
138 85.7 115.5 173.3#



#48
Acenaphthylene
Concen: 44.15 ng
RT: 9.27 min Scan# 672
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion: 152 Resp: 544667
Ion Ratio Lower Upper
152 100
151 20.0 14.6 22.0
153 12.9 10.3 15.5





#49

Dimethylphthalate

Concen: 66.69 ng

RT: 9.16 min Scan# 663

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:163 Resp: 542367

Ion Ratio Lower Upper

163 100

194 3.3 4.4 6.6#




164 10.5 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 43.61 ng

RT: 9.22 min Scan# 668

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:165 Resp: 80489

Ion Ratio Lower Upper

165 100

63 40.6 32.3 48.5



89 43.1 28.6 43.0#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0









#51
Acenaphthene
Concen: 45.34 ng
RT: 9.44 min Scan# 687
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion:154 Resp: 318212
Ion Ratio Lower Upper
154 100
153 114.9 82.1 123.1
152 55.1 37.9 56.9



#52
3-Nitroaniline
Concen: 26.40 ng
RT: 9.38 min Scan# 682
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:138 Resp: 55481
Ion Ratio Lower Upper
138 100
108 9.7 5.7 8.5#
92 107.2 67.7 101.5#









#53
 2,4-Dinitrophenol
 Concen: 21.30 ng
 RT: 9.48 min Scan# 691
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:184 Resp: 16561
 Ion Ratio Lower Upper
 184 100
 63 147.5 35.4 53.2#
 154 83.7 32.2 48.4#



#54
 Dibenzofuran
 Concen: 45.80 ng
 RT: 9.61 min Scan# 702
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:168 Resp: 469387
 Ion Ratio Lower Upper
 168 100
 139 38.4 24.2 36.2#
 169 13.2 10.6 15.8









#55
4-Nitrophenol
Concen: 226.88 ng
RT: 9.61 min Scan# 702
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion	Ratio	Lower	Upper
139	100		
109	1.9	12.7	52.7#
65	6.0	45.9	85.9#



#56
2,4-Dinitrotoluene
Concen: 40.64 ng
RT: 9.61 min Scan# 702
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
165	100		
63	102.7	29.8	44.6#
89	118.9	44.5	66.7#
182	2.1	3.0	4.4#









#57
 Fluorene
 Concen: 41.95 ng
 RT: 9.94 min Scan# 731
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion	Ratio	Lower	Upper
166	100		
165	95.4	72.6	109.0
167	14.1	10.6	16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 48.86 ng
 RT: 9.74 min Scan# 713
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
232	100		
131	55.9	38.5	57.7
130	2.4	1.8	2.6
166	33.4	25.0	37.6





#59

Diethylphthalate

Concen: 42.46 ng

RT: 9.85 min Scan# 723

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:149 Resp: 363188

Ion Ratio Lower Upper

149 100

177 17.9 17.5 26.3




150 12.3 9.4 14.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 47.19 ng

RT: 9.95 min Scan# 732

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:204 Resp: 171060

Ion Ratio Lower Upper

204 100

206 33.2 24.8 37.2



141 55.1 42.2 63.4




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 37.27 ng

RT: 9.99 min Scan# 735

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:138 Resp: 79124

Ion Ratio Lower Upper

138 100

92 60.3 12.6 52.6#




108 88.1 12.4 52.4#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 42.40 ng

RT: 10.10 min Scan# 745

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Tgt Ion: 77 Resp: 403231


Ion Ratio Lower Upper

77 100

182 14.4 15.1 55.1#

105 20.0 3.4 43.4

51 37.1 12.0 52.0






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 10.87 min Scan# 812
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion	Ratio	Lower	Upper
188	100		
94	8.1	11.1	16.7#
80	8.6	11.0	16.4#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 39.78 ng
 RT: 10.02 min Scan# 738
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
198	100		
51	44.4	39.6	79.6
105	43.0	28.5	68.5









#65
 n-Nitrosodiphenylamine
 Concen: 42.76 ng
 RT: 10.08 min Scan# 743
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:169 Resp: 316087
 Ion Ratio Lower Upper
 169 100
 168 68.3 48.9 73.3
 167 37.3 26.2 39.4



#66
 4-Bromophenyl-phenylether
 Concen: 46.18 ng
 RT: 10.43 min Scan# 774
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:248 Resp: 94858
 Ion Ratio Lower Upper
 248 100
 250 96.8 78.5 117.7
 141 72.1 59.0 88.6









#67
Hexachlorobenzene
Concen: 45.03 ng
RT: 10.49 min Scan# 779
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD

Tot Ion	Ratio	Lower	Upper
284	100		
142	21.4	29.5	44.3#
249	30.0	19.5	29.3#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 49.76 ng
RT: 10.62 min Scan# 790
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.1	13.2	53.2
215	45.5	41.8	81.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 84.12 ng
 RT: 10.68 min Scan# 796
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD

Tot Ion:266 Resp: 88479
 Ion Ratio Lower Upper
 266 100
 268 61.8 49.8 74.6
 264 63.7 50.2 75.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 39.71 ng
 RT: 10.89 min Scan# 814
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:178 Resp: 448465
 Ion Ratio Lower Upper
 178 100
 176 19.7 13.8 20.8
 179 15.6 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 44.55 ng
 RT: 10.95 min Scan# 819
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:178 Resp: 516938
 Ion Ratio Lower Upper
 178 100
 176 18.7 14.1 21.1
 179 15.1 12.2 18.2



#72
 Carbazole
 Concen: 43.23 ng
 RT: 11.11 min Scan# 833
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:167 Resp: 470721
 Ion Ratio Lower Upper
 167 100
 166 21.4 15.7 23.5
 139 13.9 9.5 14.3





#73

Di-n-butylphthalate

Concen: 44.49 ng

RT: 11.46 min Scan# 864

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:149 Resp: 572033

Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.2




104 5.1 2.4 3.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 43.83 ng

RT: 12.07 min Scan# 917

Delta R.T. -0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:202 Resp: 535785

Ion Ratio Lower Upper

202 100

101 5.4 0.0 38.3



203 17.5 0.0 37.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E









#75
Chrysene-d12
Concen: 20.00 ng
RT: 13.47 min Scan# 1040
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion:240 Resp: 163985
Ion Ratio Lower Upper
240 100
120 11.7 16.2 24.2#
236 25.6 18.4 27.6



#76
Benzidine
Concen: 16.68 ng
RT: 12.20 min Scan# 929
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:184 Resp: 117421
Ion Ratio Lower Upper
184 100
185 13.9 11.8 17.6
183 10.2 7.6 11.4









#77
Pvrene
Concen: 44.20 ng
RT: 12.28 min Scan# 936
Delta R.T. 0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion:202 Resp: 536937
Ion Ratio Lower Upper
202 100
200 20.2 15.0 22.4
203 18.2 14.1 21.1



#78
Terphenyl-d14
Concen: 92.92 ng
RT: 12.45 min Scan# 950
Delta R.T. -0.01 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:244 Resp: 654554
Ion Ratio Lower Upper
244 100
212 7.9 4.3 6.5#
122 14.4 17.4 26.2#





#79

Butylbenzylphthalate

Concen: 45.99 ng

RT: 12.94 min Scan# 993

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:149 Resp: 242961

Ion Ratio Lower Upper

149 100

91 78.4 48.6 72.8#




206 16.0 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 41.40 ng

RT: 13.46 min Scan# 1039

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:228 Resp: 432342

Ion Ratio Lower Upper

228 100

226 27.2 20.0 30.0



229 19.8 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 19.85 ng
 RT: 13.45 min Scan# 1038
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:252 Resp: 69917
 Ion Ratio Lower Upper
 252 100
 254 65.7 51.4 77.2
 126 7.2 16.4 24.6#



#82
 Chrysene
 Concen: 41.23 ng
 RT: 13.51 min Scan# 1043
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:228 Resp: 385924
 Ion Ratio Lower Upper
 228 100
 226 29.7 21.0 31.4
 229 19.2 15.6 23.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 44.77 ng
 RT: 13.51 min Scan# 1043
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:149 Resp: 312100
 Ion Ratio Lower Upper
 149 100
 167 26.8 24.2 36.2
 279 3.3 3.8 5.6#



#84
 Di-n-octyl phthalate
 Concen: 42.04 ng
 RT: 14.11 min Scan# 1096
 Delta R.T. -0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:149 Resp: 482624
 Ion Ratio Lower Upper
 149 100
 167 1.4 1.0 1.6
 43 13.3 10.2 15.2









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 38.45 ng
 RT: 15.86 min Scan# 1249
 Delta R.T. 0.00 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58




Instrument :
 BNA_F
 ClientSampleId :
 SB-07(7-8.5)MSD


Tot Ion:276 Resp: 264106
 Ion Ratio Lower Upper
 276 100
 138 23.6 25.7 38.5#
 277 24.4 15.6 23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 14.78 min Scan# 1154
 Delta R.T. -0.01 min
 Lab File: BF081395.D
 Acq: 4 Sep 2015 00:58

Tgt Ion:264 Resp: 107495
 Ion Ratio Lower Upper
 264 100
 260 23.8 16.7 25.1
 265 21.4 17.2 25.8









#87
Benzo(b)fluoranthene
Concen: 83.24 ng
RT: 14.48 min Scan# 1128
Delta R.T. 0.02 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58




Instrument :
BNA_F
ClientSampleId :
SB-07(7-8.5)MSD


Tot Ion:252 Resp: 638823
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 9.8 17.1 25.7#



#88
Benzo(k)fluoranthene
Concen: 100.32 ng
RT: 14.48 min Scan# 1128
Delta R.T. -0.00 min
Lab File: BF081395.D
Acq: 4 Sep 2015 00:58

Tgt Ion:252 Resp: 638823
Ion Ratio Lower Upper
252 100
253 22.2 17.0 25.4
125 9.8 16.0 24.0#





#89

Benzo(a)pyrene

Concen: 41.97 ng

RT: 14.73 min Scan# 1150

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:252 Resp: 272923

Ion Ratio Lower Upper

252 100

253 22.3 17.2 25.8




125 12.0 18.0 27.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 44.92 ng

RT: 15.87 min Scan# 1250

Delta R.T. -0.01 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58


Tgt Ion:278 Resp: 225709

Ion Ratio Lower Upper

278 100

139 14.4 26.3 39.5#



279 23.3 18.2 27.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 46.02 ng

RT: 16.18 min Scan# 1277

Delta R.T. 0.00 min

Lab File: BF081395.D

Acq: 4 Sep 2015 00:58

Instrument :

BNA_F

ClientSampleId :

SB-07(7-8.5)MSD


Tot Ion:276 Resp: 220711

Ion Ratio Lower Upper

276 100

277 23.2 17.8 26.6

138 18.0 34.6 51.8#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

