

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090321\
 Data File : BF125358.D
 Acq On : 3 Sep 2021 21:26
 Operator : JU/CG
 Sample : M3646-15 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 99-01

Quant Time: Sep 04 00:57:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 03 11:56:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.898	152	133453	20.000	ng	0.00
21) Naphthalene-d8	8.180	136	436898	20.000	ng	# 0.00
39) Acenaphthene-d10	9.974	164	154932	20.000	ng	0.04
64) Phenanthrene-d10	11.527	188	184889	20.000	ng	# 0.10
76) Chrysene-d12	14.127	240	282400	20.000	ng	0.06
86) Perylene-d12	15.598	264	298421	20.000	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	80024	9.076	ng	-0.01
7) Phenol-d6	6.586	99	22615	1.983	ng	0.05
23) Nitrobenzene-d5	7.457	82	141144	16.264	ng	-0.01
42) 2,4,6-Tribromophenol	10.780	330	63397	40.992	ng	0.05
45) 2-Fluorobiphenyl	9.269	172	280244	25.210	ng	0.01
79) Terphenyl-d14	13.098	244	277800	15.671	ng	0.09
Target Compounds						Qvalue
31) Naphthalene	8.204	128	211382	9.447	ng	100
35) Caprolactam	8.604	113	22740	13.023	ng	# 1
37) 2-Methylnaphthalene	8.898	142	264437	17.885	ng	97
38) 1-Methylnaphthalene	8.998	142	169106	12.248	ng	95
46) 1,1'-Biphenyl	9.374	154	51086	4.090	ng	76
48) 2-Nitroaniline	9.492	65	21519	6.839	ng	# 57
49) Acenaphthylene	9.827	152	51430	3.477	ng	# 1
50) Dimethylphthalate	9.663	163	57758	5.380	ng	# 1
51) 2,6-Dinitrotoluene	9.757	165	67303	28.909	ng	# 49
53) 3-Nitroaniline	9.927	138	89783	35.761	ng	# 33
54) 2,4-Dinitrophenol	9.898	184	2037	2.215	ng	# 1
55) Dibenzofuran	10.180	168	51758	3.922	ng	# 1
56) 4-Nitrophenol	10.086	139	23675	11.908	ng	# 1
57) 2,4-Dinitrotoluene	10.163	165	27383	9.718	ng	# 42
58) Fluorene	10.539	166	79361	8.134	ng	# 1
60) Diethylphthalate	10.433	149	46875	4.478	ng	# 8
61) 4-Chlorophenyl-phenyle...	10.557	204	44994	9.121	ng	# 1
62) 4-Nitroaniline	10.410	138	58706	26.003	ng	92
65) 4,6-Dinitro-2-methylph...	10.604	198	4445	4.542	ng	# 1
66) n-Nitrosodiphenylamine	10.657	169	214644	32.422	ng	# 53
69) Atrazine	11.198	200	14416	7.589	ng	# 1
70) Pentachlorophenol	11.363	266	32018	24.008	ng	# 26
71) Phenanthrene	11.551	178	380006	37.301	ng	# 16
72) Anthracene	11.651	178	79493	7.917	ng	# 1
73) Carbazole	11.815	167	22133	2.412	ng	# 1
74) Di-n-butylphthalate	12.086	149	72339	6.643	ng	# 29
75) Fluoranthene	12.751	202	35481	3.510	ng	# 1
84) Bis(2-ethylhexyl)phtha...	14.074	149	70106	5.359	ng	# 70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090321\
Data File : BF125358.D
Acq On : 3 Sep 2021 21:26
Operator : JU/CG
Sample : M3646-15 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
99-01

Quant Time: Sep 04 00:57:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 03 11:56:34 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.898 min Scan# 818
Delta R.T. 0.000 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:152 Resp: 133453
Ion Ratio Lower Upper
152 100
150 159.7 117.6 176.4
115 60.2 47.1 70.7



#5
2-Fluorophenol
Concen: 9.076 ng
RT: 5.510 min Scan# 582
Delta R.T. -0.012 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:112 Resp: 80024
Ion Ratio Lower Upper
112 100
64 68.7 44.2 66.2#
63 34.7 26.2 39.2









#7
Phenol-d6
Concen: 1.983 ng
RT: 6.586 min Scan# 765
Delta R.T. 0.053 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion: 99 Resp: 22615
Ion Ratio Lower Upper
99 100
42 29.8 23.5 35.3
71 25.8 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.180 min Scan# 1036
Delta R.T. -0.006 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:136 Resp: 436898
Ion Ratio Lower Upper
136 100
137 12.1 9.8 14.6
54 13.9 6.4 9.6#
68 9.5 3.3 4.9#









#23
Nitrobenzene-d5
Concen: 16.264 ng
RT: 7.457 min Scan# 913
Delta R.T. -0.012 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion: 82 Resp: 141144
Ion Ratio Lower Upper
82 100
128 39.9 40.6 60.8#
54 61.4 49.6 74.4



#31
Naphthalene
Concen: 9.447 ng
RT: 8.204 min Scan# 1040
Delta R.T. -0.000 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion: 128 Resp: 211382
Ion Ratio Lower Upper
128 100
129 11.8 9.4 14.0
127 13.3 10.9 16.3









#35
Caprolactam
Concen: 13.023 ng
RT: 8.604 min Scan# 1108
Delta R.T. -0.024 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampleId :
99-01


Tgt Ion:113 Resp: 22740
Ion Ratio Lower Upper
113 100
55 1512.7 154.1 194.1#
56 586.2 108.7 148.7#



#37
2-Methylnaphthalene
Concen: 17.885 ng
RT: 8.898 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:142 Resp: 264437
Ion Ratio Lower Upper
142 100
141 86.4 71.2 106.8









#38
1-Methylnaphthalene
Concen: 12.248 ng
RT: 8.998 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:142 Resp: 169106
Ion Ratio Lower Upper
142 100
141 89.7 76.0 114.0



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.974 min Scan# 1341
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:164 Resp: 154932
Ion Ratio Lower Upper
164 100
162 86.2 81.3 121.9
160 43.1 38.0 57.0









#42
2,4,6-Tribromophenol
Concen: 40.992 ng
RT: 10.780 min Scan# 1478
Delta R.T. 0.047 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:330 Resp: 63397
Ion Ratio Lower Upper
330 100
332 97.6 76.2 114.4
141 0.0 44.6 66.8#



#45
2-Fluorobiphenyl
Concen: 25.210 ng
RT: 9.269 min Scan# 1221
Delta R.T. 0.012 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:172 Resp: 280244
Ion Ratio Lower Upper
172 100
171 34.5 26.5 39.7
170 23.0 18.8 28.2









#46
1,1'-Biphenyl
Concen: 4.090 ng
RT: 9.374 min Scan# 1239
Delta R.T. 0.012 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:154 Resp: 51086
Ion Ratio Lower Upper
154 100
153 56.5 20.0 60.0
76 15.3 0.0 30.3



#48
2-Nitroaniline
Concen: 6.839 ng
RT: 9.492 min Scan# 1259
Delta R.T. 0.012 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion: 65 Resp: 21519
Ion Ratio Lower Upper
65 100
92 84.4 55.5 83.3#
138 53.4 93.5 140.3#





#49

Acenaphthylene

Concen: 3.477 ng

RT: 9.827 min Scan# 1316

Delta R.T. 0.023 min

Lab File: BF125358.D

Acq: 3 Sep 2021 21:26





Instrument :

BNA_F

ClientSampleId :

99-01

Tgt Ion:152 Resp: 51430
 Ion Ratio Lower Upper
 152 100
 151 149.0 15.7 23.5#
 153 70.4 9.8 14.8#



#50

Dimethylphthalate

Concen: 5.380 ng




RT: 9.663 min Scan# 1288


Delta R.T. -0.000 min

Lab File: BF125358.D

Acq: 3 Sep 2021 21:26

Tgt Ion:163 Resp: 57758
 Ion Ratio Lower Upper
 163 100
 194 120.1 2.6 3.8#
 164 34.4 7.4 11.0#









#51
2,6-Dinitrotoluene
Concen: 28.909 ng
RT: 9.757 min Scan# 1304
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:165 Resp: 67303
Ion Ratio Lower Upper
165 100
63 22.3 42.6 63.8#
89 10.3 40.6 60.8#



#53
3-Nitroaniline
Concen: 35.761 ng
RT: 9.927 min Scan# 1333
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:138 Resp: 89783
Ion Ratio Lower Upper
138 100
108 91.2 9.3 13.9#
92 47.1 76.5 114.7#









#54
2,4-Dinitrophenol
Concen: 2.215 ng
RT: 9.898 min Scan# 1328
Delta R.T. -0.100 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:184 Resp: 2037
Ion Ratio Lower Upper
184 100
63 1010.9 59.3 88.9#
154 1312.4 68.6 102.8#



#55
Dibenzofuran
Concen: 3.922 ng
RT: 10.180 min Scan# 1376
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:168 Resp: 51758
Ion Ratio Lower Upper
168 100
139 191.0 32.7 49.1#
169 195.4 11.2 16.8#









#56
4-Nitrophenol
Concen: 11.908 ng
RT: 10.086 min Scan# 1360
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampleId :
99-01


Tgt Ion:139 Resp: 23675
Ion Ratio Lower Upper
139 100
109 609.0 60.4 100.4#
65 97.5 58.4 98.4



#57
2,4-Dinitrotoluene
Concen: 9.718 ng
RT: 10.163 min Scan# 1373
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:165 Resp: 27383
Ion Ratio Lower Upper
165 100
63 23.2 40.9 61.3#
89 14.2 59.7 89.5#
182 25.1 0.0 0.0#









#58
Fluorene
Concen: 8.134 ng
RT: 10.539 min Scan# 1437
Delta R.T. 0.053 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:166 Resp: 79361
Ion Ratio Lower Upper
166 100
165 206.4 78.1 117.1#
167 87.9 11.1 16.7#



#60
Diethylphthalate
Concen: 4.478 ng
RT: 10.433 min Scan# 1419
Delta R.T. 0.076 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:149 Resp: 46875
Ion Ratio Lower Upper
149 100
177 72.2 16.5 24.7#
150 36.4 10.1 15.1#









#61
4-Chlorophenyl-phenylether
Concen: 9.121 ng
RT: 10.557 min Scan# 1440
Delta R.T. 0.076 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:204 Resp: 44994
Ion Ratio Lower Upper
204 100
206 97.1 27.6 41.4#
141 265.9 58.6 88.0#



#62
4-Nitroaniline
Concen: 26.003 ng
RT: 10.410 min Scan# 1415
Delta R.T. -0.100 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:138 Resp: 58706
Ion Ratio Lower Upper
138 100
92 46.2 29.6 69.6
108 88.9 78.1 118.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.527 min Scan# 1605
Delta R.T. 0.100 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:188 Resp: 184889
Ion Ratio Lower Upper
188 100
94 83.2 6.7 10.1#
80 73.3 7.5 11.3#



#65
4,6-Dinitro-2-methylphenol
Concen: 4.542 ng
RT: 10.604 min Scan# 1448
Delta R.T. 0.065 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:198 Resp: 4445
Ion Ratio Lower Upper
198 100
51 245.3 45.3 85.3#
105 1592.7 30.6 70.6#









#66
n-Nitrosodiphenylamine
Concen: 32.422 ng
RT: 10.657 min Scan# 1457
Delta R.T. 0.059 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampleId :
99-01


Tgt Ion:169 Resp: 214644
Ion Ratio Lower Upper
169 100
168 16.9 50.8 76.2#
167 22.0 28.8 43.2#



#69
Atrazine
Concen: 7.589 ng
RT: 11.198 min Scan# 1549
Delta R.T. 0.076 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:200 Resp: 14416
Ion Ratio Lower Upper
200 100
173 316.0 11.1 51.1#
215 368.7 22.7 62.7#









#70
Pentachlorophenol
Concen: 24.008 ng
RT: 11.363 min Scan# 1577
Delta R.T. 0.129 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:266 Resp: 32018
Ion Ratio Lower Upper
266 100
268 10.4 48.9 73.3#
264 136.0 55.9 83.9#



#71
Phenanthrene
Concen: 37.301 ng
RT: 11.551 min Scan# 1609
Delta R.T. 0.100 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:178 Resp: 380006
Ion Ratio Lower Upper
178 100
176 27.4 14.7 22.1#
179 82.0 11.5 17.3#





#72

Anthracene

Concen: 7.917 ng

RT: 11.651 min Scan# 1626

Delta R.T. 0.147 min

Lab File: BF125358.D

Acq: 3 Sep 2021 21:26





Instrument :

BNA_F

ClientSampleId :

99-01

Tgt Ion:178 Resp: 79493
 Ion Ratio Lower Upper
 178 100
 176 37.8 14.3 21.5#
 179 208.4 12.2 18.4#



#73

Carbazole

Concen: 2.412 ng




RT: 11.815 min Scan# 1654


Delta R.T. 0.159 min

Lab File: BF125358.D

Acq: 3 Sep 2021 21:26

Tgt Ion:167 Resp: 22133
 Ion Ratio Lower Upper
 167 100
 166 83.9 17.4 26.0#
 139 134.1 11.0 16.4#









#74
Di-n-butylphthalate
Concen: 6.643 ng
RT: 12.086 min Scan# 1700
Delta R.T. 0.106 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:149 Resp: 72339
Ion Ratio Lower Upper
149 100
150 44.9 7.4 11.2#
104 12.9 4.1 6.1#



#75
Fluoranthene
Concen: 3.510 ng
RT: 12.751 min Scan# 1813
Delta R.T. 0.112 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:202 Resp: 35481
Ion Ratio Lower Upper
202 100
101 17.0 0.0 30.7
203 118.6 0.0 36.3#









#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2047
Delta R.T. 0.059 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26




Instrument :
BNA_F
ClientSampled :
99-01


Tgt Ion:240 Resp: 282400
Ion Ratio Lower Upper
240 100
120 15.4 10.8 16.2
236 27.9 21.6 32.4



#79
Terphenyl-d14
Concen: 15.671 ng
RT: 13.098 min Scan# 1872
Delta R.T. 0.088 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:244 Resp: 277800
Ion Ratio Lower Upper
244 100
212 9.2 5.9 8.9#
122 19.5 9.0 13.4#









#84
Bis(2-ethylhexyl)phthalate
Concen: 5.359 ng
RT: 14.074 min Scan# 2038
Delta R.T. 0.035 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Instrument :
BNA_F
ClientSampled :
99-01

Tgt Ion:149 Resp: 70106
Ion Ratio Lower Upper
149 100
167 42.3 21.4 32.0#
279 11.3 2.6 4.0#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.598 min Scan# 2297
Delta R.T. 0.041 min
Lab File: BF125358.D
Acq: 3 Sep 2021 21:26

Tgt Ion:264 Resp: 298421
Ion Ratio Lower Upper
264 100
260 23.1 19.6 29.4
265 21.9 17.2 25.8

