

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090418\
 Data File : BF108769.D
 Acq On : 5 Sep 2018 6:29
 Operator : JU/SJ
 Sample : J4741-06 10X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AOC2-04-B

Manual Integrations
 APPROVED

Sohil
 9/5/2018 12:48:12 PM

Quant Time: Sep 05 08:39:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 30 15:06:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.00	152	43035	20.00	ng	-0.01
21) Naphthalene-d8	8.28	136	215324	20.00	ng	0.00
38) Acenaphthene-d10	10.03	164	109731	20.00	ng	-0.02
63) Phenanthrene-d10	11.53	188	245033	20.00	ng	-0.02
75) Chrysene-d12	14.19	240	210291	20.00	ng	-0.01
86) Perylene-d12	15.74	264	154030	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.73	112	120	0.05	ng	0.08
7) Phenol-d6	6.81	99	20391m	5.75	ng	0.16
23) Nitrobenzene-d5	7.68	82	15214m	3.59	ng	0.11
41) 2,4,6-Tribromophenol	10.84	330	6576	4.54	ng	-0.01
44) 2-Fluorobiphenyl	9.35	172	34980	4.29	ng	-0.01
78) Terphenyl-d14	13.11	244	55283	5.11	ng	-0.02
Target Compounds						
74) Fluoranthene	12.75	202	33066	2.375	ng	Qvalue 96
87) Benzo(b)fluoranthene	15.28	252	19016m	2.010	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF090418\
Data File : BF108769.D
Acq On : 5 Sep 2018 6:29
Operator : JU/SJ
Sample : J4741-06 10X
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AOC2-04-B

Quant Time: Sep 05 08:39:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF083018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 30 15:06:55 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
9/5/2018 12:48:12 PM

