


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090418\
 Data File : BF108759.D
 Acq On : 5 Sep 2018 1:56
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF083018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 30 15:06:55 2018



AutoFind: Scans 1600, 1601, 1602; Background Corrected with Scan 1595

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.4	14840	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.4	12911	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.6	16484	PASS
197	198	0.00	2	0.6	245	PASS
198	198	100	100	100.0	38655	PASS
199	198	5	9	6.7	2574	PASS
275	198	10	60	29.7	11480	PASS
365	198	1	100	3.7	1414	PASS
441	198	0.01	100	14.6	5641	PASS
442	198	50	100	91.2	35264	PASS
443	442	15	24	19.7	6958	PASS