

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090515\
 Data File : BF081500.D
 Acq On : 6 Sep 2015 10:12
 Operator : UM/IZ
 Sample : G3498-01DL 25X
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP206BR-29-31DL

Manual Integrations
 APPROVED

MMDadoda
 9/8/2015 3:57:26 PM

Quant Time: Sep 07 09:02:03 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 07 02:47:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.35	152	49526	20.00	ng	-0.01
21) Naphthalene-d8	7.64	136	186749	20.00	ng	-0.01
38) Acenaphthene-d10	9.38	164	85824	20.00	ng	-0.01
63) Phenanthrene-d10	10.84	188	164525	20.00	ng	-0.01
75) Chrysene-d12	13.45	240	113422	20.00	ng	-0.01
86) Perylene-d12	14.77	264	84796	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.90	112	17702	5.55	ng	0.01
7) Phenol-d6	6.02	99	29368	6.95	ng	-0.01
23) Nitrobenzene-d5	6.94	82	14552	3.76	ng	-0.01
41) 2,4,6-Tribromophenol	10.17	330	3616	4.73	ng	-0.01
44) 2-Fluorobiphenyl	8.73	172	28574	4.27	ng	0.00
78) Terphenyl-d14	12.43	244	21721	4.46	ng	0.00
Target Compounds						
31) Naphthalene	7.67	128	493522	49.55	ng	97
37) 2-Methylnaphthalene	8.36	142	146451	22.60	ng	# 91
45) 1,1'-Biphenyl	8.82	154	18260	2.31	ng	94
48) Acenaphthylene	9.24	152	125784	12.43	ng	97
57) Fluorene	9.93	166	47586m	7.46	ng	
70) Phenanthrene	10.87	178	160698	17.57	ng	97
71) Anthracene	10.92	178	45176	4.81	ng	99
74) Fluoranthene	12.04	202	46232	4.67	ng	87
77) Pyrene	12.26	202	68959	8.21	ng	99
80) Benzo(a)anthracene	13.44	228	16238	2.25	ng	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090515\
Data File : BF081500.D
Acq On : 6 Sep 2015 10:12
Operator : UM/IZ
Sample : G3498-01DL 25X
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MGP206BR-29-31DL

Manual Integrations
APPROVED

MMDadoda
9/8/2015 3:57:26 PM

Quant Time: Sep 07 09:02:03 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 07 02:47:26 2015
Response via : Initial Calibration

