

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090616\
 Data File : BF089990.D
 Acq On : 6 Sep 2016 15:09
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 06 17:39:01 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 13:29:08 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	-0.02
2	1,4-Dioxane	0.592	0.582	1.7	117	-0.03
3	Pyridine	1.579	1.486	5.9	102	-0.03
4	n-Nitrosodimethylamine	0.779	0.719	7.7	101	-0.05
5 S	2-Fluorophenol	1.250	1.265	-1.2	109	-0.02
6	Aniline	2.072	1.957	5.6	112	-0.01
7 S	Phenol-d6	1.518	1.463	3.6	107	-0.01
8	2-Chlorophenol	1.356	1.341	1.1	110	-0.02
9	Benzaldehyde	0.980	0.937	4.4	107	-0.02
10 C	Phenol	1.766	1.616	8.5	106	-0.01
11	bis(2-Chloroethyl)ether	1.336	1.371	-2.6	122	-0.02
12	1,3-Dichlorobenzene	1.514	1.546	-2.1	118	-0.02
13 C	1,4-Dichlorobenzene	1.549	1.580	-2.0	120	-0.02
14	1,2-Dichlorobenzene	1.389	1.343	3.3	106	-0.02
15	Benzyl Alcohol	0.951	0.984	-3.5	116	-0.01
16	2,2'-oxybis(1-Chloropropane	2.547	2.429	4.6	105	-0.02
17	2-Methylphenol	1.083	1.105	-2.0	115	-0.01
18	Hexachloroethane	0.531	0.520	2.1	111	-0.02
19 P	n-Nitroso-di-n-propylamine	0.989	0.901	8.9	98	-0.02
20	3+4-Methylphenols	1.312	1.398	-6.6	135	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	117	-0.02
22	Acetophenone	0.449	0.423	5.8	114	-0.02
23 S	Nitrobenzene-d5	0.345	0.329	4.6	112	-0.02
24	Nitrobenzene	0.367	0.343	6.5	118	-0.02
25	Isophorone	0.701	0.654	6.7	113	-0.02
26 C	2-Nitrophenol	0.175	0.197	-12.6	132	-0.01
27	2,4-Dimethylphenol	0.324	0.308	4.9	123	-0.02
28	bis(2-Chloroethoxy)methane	0.423	0.405	4.3	109	-0.01
29 C	2,4-Dichlorophenol	0.291	0.287	1.4	123	-0.02
30	1,2,4-Trichlorobenzene	0.318	0.304	4.4	109	-0.02
31	Naphthalene	0.996	0.949	4.7	122	-0.02
32	Benzoic acid	0.218	0.257	-17.9	140	0.00
33	4-Chloroaniline	0.402	0.414	-3.0	119	-0.01
34 C	Hexachlorobutadiene	0.186	0.190	-2.2	124	-0.02
35	Caprolactam	0.092	0.083	9.8	102	-0.02
36 C	4-Chloro-3-methylphenol	0.308	0.304	1.3	117	-0.01
37	2-Methylnaphthalene	0.658	0.590	10.3	104	-0.02
38 I	Acenaphthene-d10	1.000	1.000	0.0	111	-0.01
39	1,2,4,5-Tetrachlorobenzene	0.586	0.616	-5.1	131	-0.01
40 P	Hexachlorocyclopentadiene	0.300	0.360	-20.0	129	-0.02
41 S	2,4,6-Tribromophenol	0.197	0.189	4.1	105	-0.02
42 C	2,4,6-Trichlorophenol	0.406	0.436	-7.4	112	-0.01
43	2,4,5-Trichlorophenol	0.390	0.420	-7.7	128	-0.02
44 S	2-Fluorobiphenyl	1.218	1.182	3.0	111	-0.02

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090616\
 Data File : BF089990.D
 Acq On : 6 Sep 2016 15:09
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 06 17:39:01 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 13:29:08 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.610	1.556	3.4	118	-0.02
46	2-Chloronaphthalene	1.139	1.203	-5.6	122	-0.01
47	2-Nitroaniline	0.372	0.381	-2.4	117	-0.01
48	Acenaphthylene	1.884	1.817	3.6	110	-0.01
49	Dimethylphthalate	1.446	1.294	10.5	112	-0.02
50	2,6-Dinitrotoluene	0.322	0.329	-2.2	117	-0.01
51 C	Acenaphthene	1.193	1.241	-4.0	121	-0.01
52	3-Nitroaniline	0.367	0.393	-7.1	132	-0.01
53 P	2,4-Dinitrophenol	0.130	0.177	-36.2#	136	-0.01
54	Dibenzofuran	1.599	1.507	5.8	108	-0.01
55 P	4-Nitrophenol	0.277	0.316	-14.1	117	0.00
56	2,4-Dinitrotoluene	0.408	0.442	-8.3	107	-0.01
57	Fluorene	1.169	1.015	13.2	92	-0.02
58	2,3,4,6-Tetrachlorophenol	0.334	0.326	2.4	113	-0.02
59	Diethylphthalate	1.360	1.368	-0.6	118	-0.01
60	4-Chlorophenyl-phenylether	0.550	0.511	7.1	102	-0.02
61	4-Nitroaniline	0.360	0.332	7.8	94	-0.01
62	Azobenzene	1.358	1.319	2.9	116	-0.02
63 I	Phenanthrene-d10	1.000	1.000	0.0	106	-0.02
64	4,6-Dinitro-2-methylphenol	0.128	0.124	3.1	112	-0.02
65 c	n-Nitrosodiphenylamine	0.601	0.640	-6.5	120	-0.01
66	4-Bromophenyl-phenylether	0.201	0.206	-2.5	112	-0.01
67	Hexachlorobenzene	0.230	0.245	-6.5	127	-0.02
68	Atrazine	0.201	0.182	9.5	104	-0.02
69 C	Pentachlorophenol	0.124	0.151	-21.8#	116	-0.01
70	Phenanthrene	0.996	0.905	9.1	98	-0.02
71	Anthracene	1.010	1.020	-1.0	121	-0.02
72	Carbazole	0.947	0.991	-4.6	116	-0.01
73	Di-n-butylphthalate	1.104	1.092	1.1	112	-0.02
74 C	Fluoranthene	1.044	0.922	11.7	92	-0.02
75 I	Chrysene-d12	1.000	1.000	0.0	107	-0.02
76	Benzidine	0.758	0.754	0.5	106	-0.01
77	Pyrene	1.410	1.336	5.2	112	-0.02
78 S	Terphenyl-d14	0.810	0.776	4.2	121	-0.02
79	Butylbenzylphthalate	0.569	0.586	-3.0	113	-0.01
80	Benzo(a)anthracene	1.225	1.160	5.3	104	-0.02
81	3,3'-Dichlorobenzidine	0.442	0.487	-10.2	121	-0.02
82	Chrysene	1.144	1.033	9.7	102	-0.02
83	Bis(2-ethylhexyl)phthalate	0.787	0.782	0.6	106	-0.02
84 c	Di-n-octyl phthalate	1.270	1.364	-7.4	107	-0.02
85	Indeno(1,2,3-cd)pyrene	0.846	0.882	-4.3	112	-0.03
86 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
87	Benzo(b)fluoranthene	1.253	1.224	2.3	95	-0.02

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090616\
Data File : BF089990.D
Acq On : 6 Sep 2016 15:09
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 06 17:39:01 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF083116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 02 13:29:08 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.063	0.987	7.1	120	-0.02
89 C	Benzo(a)pyrene	1.073	1.099	-2.4	111	-0.02
90	Dibenzo(a,h)anthracene	0.787	0.793	-0.8	110	-0.05
91	Benzo(g,h,i)perylene	0.792	0.788	0.5	111	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 1