

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090717\
 Data File : BF098316.D
 Acq On : 7 Sep 2017 16:25
 Operator : SJ/JU
 Sample : I5075-10
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 292

Manual Integrations
 APPROVED

Sohil
 9/8/2017 5:45:14 PM

Quant Time: Sep 07 17:34:58 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 17:28:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	111061	20.00	ng	-0.02
21) Naphthalene-d8	7.97	136	442622	20.00	ng	-0.02
38) Acenaphthene-d10	9.72	164	190189	20.00	ng	-0.02
63) Phenanthrene-d10	11.20	188	320059	20.00	ng	-0.02
75) Chrysene-d12	13.84	240	236481	20.00	ng	-0.02
86) Perylene-d12	15.26	264	221111	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	734792	100.64	ng	0.00
7) Phenol-d6	6.33	99	1008466	101.65	ng	-0.02
23) Nitrobenzene-d5	7.26	82	624492	76.65	ng	-0.02
41) 2,4,6-Tribromophenol	10.52	330	175318	106.24	ng	-0.02
44) 2-Fluorobiphenyl	9.05	172	940721	72.67	ng	-0.02
78) Terphenyl-d14	12.79	244	634483	55.95	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.45	163	164497	11.832	ng	99
70) Phenanthrene	11.23	178	176320	10.338	ng	97
71) Anthracene	11.28	178	45609	2.637	ng	99
74) Fluoranthene	12.42	202	312268	17.542	ng	97
77) Pyrene	12.64	202	292146	15.105	ng	98
80) Benzo(a)anthracene	13.83	228	151820	10.712	ng	98
82) Chrysene	13.87	228	145479	10.318	ng	98
83) Bis(2-ethylhexyl)phthalate	13.83	149	21333	2.596	ng	# 98
87) Benzo(b)fluoranthene	14.86	252	179794m	12.900	ng	
88) Benzo(k)fluoranthene	14.87	252	64393m	4.918	ng	
89) Benzo(a)pyrene	15.19	252	125788	10.195	ng	# 95
91) Benzo(a,h,i)perylene	17.02	276	67190	6.411	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090717\
Data File : BF098316.D
Acq On : 7 Sep 2017 16:25
Operator : SJ/JU
Sample : I5075-10
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
292

Manual Integrations
APPROVED

Sohil
9/8/2017 5:45:14 PM

Quant Time: Sep 07 17:34:58 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 17:28:20 2017
Response via : Initial Calibration

