

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF090717\
 Data File : BF098339.D
 Acq On : 8 Sep 2017 3:42
 Operator : SJ/JU
 Sample : I5131-02
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 090517-TIDO-F-D-S

Integration Parameters: rteint.p

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M

Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.857	469	471	481	rBV	74533	87580	2.96%	0.448%
2	5.292	541	545	555	rBV	1029867	1005321	33.94%	5.139%
3	6.328	717	721	733	rBV2	828015	732551	24.73%	3.744%
4	6.457	739	743	746	rBV	1905588	1671323	56.43%	8.543%
5	6.686	779	782	786	rBV	701106	614591	20.75%	3.142%
6	6.845	805	809	812	rBV	2340103	2188557	73.89%	11.187%
7	6.869	812	813	821	rVB	52513	29899	1.01%	0.153%
8	7.257	874	879	882	rBV	1732998	1764767	59.59%	9.021%
9	7.975	997	1001	1003	rBV	905347	883480	29.83%	4.516%
10	7.992	1003	1004	1008	rVB	675356	449658	15.18%	2.298%
11	8.686	1118	1122	1125	rBV	170703	130403	4.40%	0.667%
12	8.780	1135	1138	1142	rVB	76246	71214	2.40%	0.364%
13	9.051	1179	1184	1187	rBV	3114080	2961720	100.00%	15.139%
14	9.145	1198	1200	1206	rVB	53889	47005	1.59%	0.240%
15	9.727	1293	1299	1301	rBV	762844	748408	25.27%	3.826%
16	9.757	1301	1304	1307	rVB	291840	225756	7.62%	1.154%
17	9.927	1330	1333	1338	rBV	160040	130761	4.42%	0.668%
18	10.269	1388	1391	1397	rVB	182034	157513	5.32%	0.805%
19	10.516	1429	1433	1444	rBV	995575	994131	33.57%	5.082%
20	11.210	1547	1551	1553	rBV	741688	716260	24.18%	3.661%
21	11.233	1553	1555	1558	rVB	413073	324792	10.97%	1.660%
22	11.439	1585	1590	1598	rBB	39889	34229	1.16%	0.175%
23	11.821	1652	1655	1657	rBV2	33934	30866	1.04%	0.158%
24	12.415	1753	1756	1760	rBV	141068	113197	3.82%	0.579%
25	12.615	1787	1790	1792	rBV	51314	46418	1.57%	0.237%
26	12.645	1792	1795	1798	rVB	82954	70764	2.39%	0.362%
27	12.798	1816	1821	1824	rBV	2146416	2227764	75.22%	11.387%
28	13.321	1906	1910	1913	rBV	36349	29641	1.00%	0.152%
29	13.839	1994	1998	2002	rBV	680868	643925	21.74%	3.291%
30	15.251	2233	2238	2247	rBV	370868	430990	14.55%	2.203%


Sum of corrected areas: 19563484

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090717\
Data File : BF098339.D
Acq On : 8 Sep 2017 3:42
Operator : SJ/JU
Sample : I5131-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
090517-TIDO-F-D-S

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090717\
Data File : BF098339.D
Acq On : 8 Sep 2017 3:42
Operator : SJ/JU
Sample : I5131-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
090517-TIDO-F-D-S










Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.86	2.85 ng	87580	1,4-Dichlorobenzene-d4	6.69

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	56
2			2-Pentanol, 2,4-dimethyl-	116	C7H16O	000625-06-9	50
3			Acetic acid, cyano-, 1,1-dimethyl...	141	C7H11NO2	001116-98-9	17
4			Butyl aldoxime, 3-methyl-, syn-	101	C5H11NO	005780-40-5	16
5			3-Hydroxy-3-methyl-2-butanone	102	C5H10O2	000115-22-0	12



Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090717\
Data File : BF098339.D
Acq On : 8 Sep 2017 3:42
Operator : SJ/JU
Sample : I5131-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
090517-TIDO-F-D-S










Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown6.46 Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.46	54.39 ng	1671320	1,4-Dichlorobenzene-d4	6.69

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	25
2		Tranylcypromine-propionyl	189	C12H15NO	1000123-86-3	10
3		Bicyclo[4.2.0]octa-1,3,5-triene,...	132	C10H12	028749-81-7	9
4		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	9
5		5-Fluoro-2-chloropyrimidine	132	C4H2ClFN2	062802-42-0	9



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF090717\
Data File : BF098339.D
Acq On : 8 Sep 2017 3:42
Operator : SJ/JU
Sample : I5131-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
090517-TIDO-F-D-S

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	4.86	2.9	ng	87580	1	6.69	614591	20.0
unknown6.46	6.46	54.4	ng	1671320	1	6.69	614591	20.0