

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090721\
 Data File : BF125394.D
 Acq On : 7 Sep 2021 18:39
 Operator : JU/CG
 Sample : M3655-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CON-B

Quant Time: Sep 08 12:19:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 03 11:56:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.898	152	133198	20.000	ng	0.00
21) Naphthalene-d8	8.181	136	460972	20.000	ng	# 0.00
39) Acenaphthene-d10	9.933	164	213480	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	338952	20.000	ng	0.00
76) Chrysene-d12	14.063	240	334700	20.000	ng	# 0.00
86) Perylene-d12	15.557	264	358142	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	25722	2.923	ng	0.00
7) Phenol-d6	6.522	99	406575	35.723	ng	-0.01
23) Nitrobenzene-d5	7.457	82	618105	67.505	ng	-0.01
42) 2,4,6-Tribromophenol	0.000	330	0	0.000	ng	
45) 2-Fluorobiphenyl	9.251	172	1055699	68.923	ng	0.00
79) Terphenyl-d14	13.004	244	1090910	51.925	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090721\
Data File : BF125394.D
Acq On : 7 Sep 2021 18:39
Operator : JU/CG
Sample : M3655-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CON-B

Quant Time: Sep 08 12:19:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 03 11:56:34 2021
Response via : Initial Calibration

