


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090722\
Data File : BF130152.D
Acq On : 07 Sep 2022 09:38
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 07 02:05:07 2022






AutoFind: Scans 1560, 1561, 1562; Background Corrected with Scan 1555

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	83069	PASS
68	69	0.00	2	1.7	1415	PASS
69	198	0.00	100	35.6	82045	PASS
70	69	0.00	2	0.5	416	PASS
127	198	10	80	51.7	119123	PASS
197	198	0.00	2	0.3	581	PASS
198	198	100	100	100.0	230552	PASS
199	198	5	9	6.8	15565	PASS
275	198	10	60	27.2	62795	PASS
365	198	1	100	3.6	8383	PASS
441	198	0.01	100	13.4	30871	PASS
442	442	50	100	100.0	203387	PASS
443	442	15	24	18.3	37260	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090722\
 Data File : BF130152.D
 Acq On : 07 Sep 2022 09:38
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Sep 08 06:38:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 07 02:05:07 2022
 Response via : Initial Calibration



TIC: BF130152.D\data.ms

(70) Pentachlorophenol (C)

11.010min (-0.006) 35572.85 ng




response 239402

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	62.44
264.00	59.90	62.54
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090722\
Data File : BF130152.D
Acq On : 07 Sep 2022 09:38
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 08 06:38:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 07 02:05:07 2022
Response via : Initial Calibration



TIC: BF130152.D\data.ms

(77) Benzidine

12.545min (-0.000) 0.00 ng

response 1244922

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.30	14.86
183.00	11.80	11.83
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/7/2022	BNA_F	<u>BF130152.D</u>
Compound Name	Response	Retention Time
DDT	501770	13.368
DDD	4492	13.063
DDE	241	12.727
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4733	506503	0.93

Instrument :
BNA_F
ClientSampleId :
DFTPP