

Data Path : Z:\HPCHEM1\BNA F\DATA\BF090816\
 Data File : BF090027.D
 Acq On : 8 Sep 2016 16:29
 Operator : UM/SJ
 Sample : H4680-30
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BH-26-1FT

Quant Time: Sep 08 17:00:14 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 08 16:07:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.59	152	189328	20.00	ng	-0.01
21) Naphthalene-d8	7.88	136	776254	20.00	ng	-0.01
38) Acenaphthene-d10	9.63	164	383692	20.00	ng	0.00
63) Phenanthrene-d10	11.10	188	678725	20.00	ng	-0.01
75) Chrysene-d12	13.71	240	473552	20.00	ng	-0.01
86) Perylene-d12	15.05	264	426899	20.00	ng	-0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	1711885	144.73	ng	0.02
7) Phenol-d6	6.25	99	1976781	137.54	ng	0.00
23) Nitrobenzene-d5	7.18	82	1313376	98.13	ng	0.00
41) 2,4,6-Tribromophenol	10.42	330	514014	135.79	ng	0.00
44) 2-Fluorobiphenyl	8.97	172	2090292	89.43	ng	0.00
78) Terphenyl-d14	12.69	244	1466199	76.48	ng	0.00


Target Compounds						Qvalue
10) Phenol	6.26	94	42136	2.52	ng	86
15) Benzyl Alcohol	6.75	79	20416	2.27	ng	# 48
19) n-Nitroso-di-n-propylamine	7.18	70	170686	18.22	ng	# 74
25) Isophorone	7.18	82	1172645	43.08	ng	# 65
36) 4-Chloro-3-methylphenol	8.47	107	221264	18.52	ng	# 11
49) Dimethylphthalate	9.37	163	1370523	49.42	ng	99
50) 2,6-Dinitrotoluene	9.37	165	15401	2.50	ng	# 35
56) 2,4-Dinitrotoluene	9.63	165	50339	6.44	ng	# 31
66) 4-Bromophenyl-phenylether	10.42	248	32747	4.79	ng	# 1
70) Phenanthrene	11.12	178	365996	10.83	ng	98
71) Anthracene	11.18	178	99320	2.90	ng	98
74) Fluoranthene	12.30	202	775391	21.89	ng	94
77) Pvrene	12.53	202	623163	18.66	ng	99
80) Benzo(a)anthracene	13.70	228	353033	12.17	ng	97
82) Chrysene	13.74	228	279518	10.32	ng	98
85) Indeno(1,2,3-cd)pyrene	16.26	276	130280	6.50	ng	96
87) Benzo(b)fluoranthene	14.70	252	492338	18.41	ng	# 97
88) Benzo(k)fluoranthene	14.70	252	492338	21.70	ng	# 98
89) Benzo(a)pvrene	15.01	252	231627	10.11	ng	99
91) Benzo(a,h,i)perylene	16.62	276	108592	6.43	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF090816\
Data File : BF090027.D
Acq On : 8 Sep 2016 16:29
Operator : UM/SJ
Sample : H4680-30
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BH-26-1FT

Quant Time: Sep 08 17:00:14 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF083116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 08 16:07:52 2016
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.59 min Scan# 438
 Delta R.T. -0.01 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29

Instrument :
 BNA_F
 ClientSampleId :
 BH-26-1FT

Tot Ion:152 Resp: 189328
 Ion Ratio Lower Upper
 152 100
 150 162.3 127.4 191.2
 115 69.4 46.3 69.5




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
 Concen: 144.73 ng
 RT: 5.19 min Scan# 315
 Delta R.T. 0.02 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29

Tgt Ion:112 Resp: 1711885
 Ion Ratio Lower Upper
 112 100
 64 52.7 41.1 61.7
 63 27.4 21.2 31.8



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF0
 Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 137.54 ng

RT: 6.25 min Scan# 408

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT





Tot Ion: 99 Resp: 1976781

Ion Ratio Lower Upper

99 100

42 23.3 18.6 27.8

71 33.1 25.8 38.6



#10

Phenol

Concen: 2.52 ng

RT: 6.26 min Scan# 409

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29




Tgt Ion: 94 Resp: 42136


Ion Ratio Lower Upper

94 100

65 42.5 23.2 63.2

66 87.4 49.2 89.2





#15

Benzyl Alcohol

Concen: 2.27 ng

RT: 6.75 min Scan# 452

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion: 79 Resp: 20416

Ion Ratio Lower Upper

79 100

108 27.6 62.1 93.1#




77 97.7 50.1 75.1#



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#19

n-Nitroso-di-n-propylamine

Concen: 18.22 ng

RT: 7.18 min Scan# 489

Delta R.T. 0.14 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Tgt Ion: 70 Resp: 170686


Ion Ratio Lower Upper

70 100

42 47.9 52.1 78.1#

101 0.0 7.4 11.2#

130 2.7 17.0 25.4#






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): BF0

Ion 130.00 (129.70 to 130.70): BF0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.88 min Scan# 551

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT

Tot Ion:136 Resp: 776254


Ion Ratio Lower Upper

136 100

137 11.0 8.3 12.5

54 9.0 4.8 7.2#

68 6.0 3.7 5.5#




Abundance Ion 136.00 (135.70 to 136.70): E



Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



Time-->



#23

Nitrobenzene-d5

Concen: 98.13 ng

RT: 7.18 min Scan# 489

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion: 82 Resp: 1313376

Ion Ratio Lower Upper

82 100

128 56.1 37.8 56.8


54 50.9 46.9 70.3



Abundance Ion 82.00 (81.70 to 82.70): BF0


Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0



Time-->





#25

Isophorone

Concen: 43.08 ng

RT: 7.18 min Scan# 489

Delta R.T. -0.26 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion: 82 Resp: 1172645

Ion Ratio Lower Upper




82 100

95 0.0 5.5 8.3#

138 0.0 14.4 21.6#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 18.52 ng

RT: 8.47 min Scan# 602

Delta R.T. 0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:107 Resp: 221264

Ion Ratio Lower Upper



107 100


144 0.0 25.8 38.8#

142 0.0 79.2 118.8#



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.63 min Scan# 704

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion:164 Resp: 383692

Ion Ratio Lower Upper

164 100

162 98.4 77.1 115.7




160 44.6 35.3 52.9



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 135.79 ng

RT: 10.42 min Scan# 773

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:330 Resp: 514014

Ion Ratio Lower Upper

330 100

332 96.6 78.6 117.8



141 28.5 23.3 34.9




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 89.43 ng
RT: 8.97 min Scan# 646
Delta R.T. 0.00 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Instrument :
BNA_F
ClientSampleId :
BH-26-1FT

Tot Ion:172 Resp: 2090292
Ion Ratio Lower Upper
172 100
171 37.0 29.4 44.2
170 25.0 20.1 30.1




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#49
Dimethylphthalate
Concen: 49.42 ng
RT: 9.37 min Scan# 681
Delta R.T. -0.01 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Tgt Ion:163 Resp: 1370523
Ion Ratio Lower Upper
163 100
194 3.9 2.7 4.1
164 10.7 9.0 13.4



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#50
 2,6-Dinitrotoluene
 Concen: 2.50 ng
 RT: 9.37 min Scan# 681
 Delta R.T. -0.07 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29

Instrument :
 BNA_F
 ClientSampleId :
 BH-26-1FT

Tot Ion:165 Resp: 15401
 Ion Ratio Lower Upper
 165 100
 63 108.3 45.0 67.4#
 89 10.6 41.5 62.3#




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 89.00 (88.70 to 89.70): BF0






#56
 2,4-Dinitrotoluene
 Concen: 6.44 ng
 RT: 9.63 min Scan# 704
 Delta R.T. -0.19 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29

Tgt Ion:165 Resp: 50339
 Ion Ratio Lower Upper
 165 100
 63 2.6 32.5 48.7#
 89 2.0 48.3 72.5#
 182 0.0 1.7 2.5#



Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 89.00 (88.70 to 89.70): BF0
 Ion 182.00 (181.70 to 182.70): E









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.10 min Scan# 832
Delta R.T. -0.01 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29




Instrument :
BNA_F
ClientSampleId :
BH-26-1FT


Tot Ion:188 Resp: 678725
Ion Ratio Lower Upper
188 100
94 11.7 5.9 8.9#
80 12.7 5.8 8.8#



#66
4-Bromophenyl-phenylether
Concen: 4.79 ng
RT: 10.42 min Scan# 773
Delta R.T. -0.24 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Tgt Ion:248 Resp: 32747
Ion Ratio Lower Upper
248 100
250 200.0 77.3 115.9#
141 265.3 55.8 83.8#









#70
Phenanthrene
Concen: 10.83 ng
RT: 11.12 min Scan# 834
Delta R.T. -0.01 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Instrument :
BNA_F
ClientSampleId :
BH-26-1FT

Tot Ion:178 Resp: 365996
Ion Ratio Lower Upper
178 100
176 19.4 16.5 24.7
179 15.7 12.8 19.2




Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E






#71
Anthracene
Concen: 2.90 ng
RT: 11.18 min Scan# 839
Delta R.T. 0.00 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Tgt Ion:178 Resp: 99320
Ion Ratio Lower Upper
178 100
176 18.3 15.8 23.8
179 15.7 13.0 19.4



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#74

Fluoranthene

Concen: 21.89 ng

RT: 12.30 min Scan# 937

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion:202 Resp: 775391

Ion Ratio Lower Upper

202 100

101 14.2 0.0 28.5




203 18.3 0.0 38.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.71 min Scan# 1061

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:240 Resp: 473552

Ion Ratio Lower Upper

240 100

120 12.1 5.7 8.5#



236 26.5 21.3 31.9




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pvrene

Concen: 18.66 ng

RT: 12.53 min Scan# 957

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion:202 Resp: 623163

Ion Ratio Lower Upper

202 100

200 21.8 17.6 26.4




203 18.6 14.6 21.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 76.48 ng

RT: 12.69 min Scan# 971

Delta R.T. 0.00 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:244 Resp: 1466199

Ion Ratio Lower Upper

244 100

212 7.7 6.1 9.1



122 11.8 9.6 14.4




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#80

Benzo(a)anthracene

Concen: 12.17 ng

RT: 13.70 min Scan# 1060

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion:228 Resp: 353033

Ion Ratio Lower Upper

228 100

226 28.8 22.2 33.2




229 21.2 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

Chrysene

Concen: 10.32 ng

RT: 13.74 min Scan# 1063

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:228 Resp: 279518

Ion Ratio Lower Upper

228 100

226 30.3 25.2 37.8



229 21.5 16.2 24.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 6.50 ng
 RT: 16.26 min Scan# 1284
 Delta R.T. -0.01 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29




Instrument :
 BNA_F
 ClientSampleId :
 BH-26-1FT


Tot Ion	Ratio	Lower	Upper
276	100		
138	25.8	22.6	34.0
277	26.4	20.1	30.1



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.05 min Scan# 1178
 Delta R.T. -0.01 min
 Lab File: BF090027.D
 Acq: 8 Sep 2016 16:29

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.4	29.0
265	21.5	17.4	26.2





#87

Benzo(b)fluoranthene

Concen: 18.41 ng

RT: 14.70 min Scan# 1147

Delta R.T. -0.01 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29

Instrument :

BNA_F

ClientSampleId :

BH-26-1FT


Tot Ion:252 Resp: 492338

Ion Ratio Lower Upper

252 100

253 22.3 18.0 27.0




125 11.7 5.7 8.5#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 21.70 ng

RT: 14.70 min Scan# 1147

Delta R.T. -0.03 min

Lab File: BF090027.D

Acq: 8 Sep 2016 16:29


Tgt Ion:252 Resp: 492338

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8



125 11.7 7.3 10.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(a)pyrene
Concen: 10.11 ng
RT: 15.01 min Scan# 1174
Delta R.T. 0.00 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Instrument :
BNA_F
ClientSampleId :
BH-26-1FT

Tot Ion:252 Resp: 231627
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 8.8 6.3 9.5




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#91
Benzo(g,h,i)perylene
Concen: 6.43 ng
RT: 16.62 min Scan# 1315
Delta R.T. -0.02 min
Lab File: BF090027.D
Acq: 8 Sep 2016 16:29

Tgt Ion:276 Resp: 108592
Ion Ratio Lower Upper
276 100
277 25.6 19.0 28.4
138 27.1 16.4 24.6#



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

