

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090816\
 Data File : BF090043.D
 Acq On : 9 Sep 2016 00:05
 Operator : UM/SJ
 Sample : H4688-03 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SS-AST3

Quant Time: Sep 09 06:36:46 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 08 16:07:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.59	152	192728	20.00	ng	-0.01
21) Naphthalene-d8	7.88	136	795189	20.00	ng	-0.01
38) Acenaphthene-d10	9.63	164	353260	20.00	ng	0.00
63) Phenanthrene-d10	11.10	188	570381	20.00	ng	-0.01
75) Chrysene-d12	13.71	240	490691	20.00	ng	-0.01
86) Perylene-d12	15.05	264	358709	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.16	112	107178	8.90	ng	0.00
7) Phenol-d6	6.24	99	157862	10.79	ng	-0.01
23) Nitrobenzene-d5	7.16	82	85885	6.26	ng	-0.01
41) 2,4,6-Tribromophenol	10.41	330	36204	10.39	ng	-0.01
44) 2-Fluorobiphenyl	8.96	172	217112	10.09	ng	-0.01
78) Terphenyl-d14	12.69	244	163490	8.23	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.36	163	86901	3.40	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF090816\
Data File : BF090043.D
Acq On : 9 Sep 2016 00:05
Operator : UM/SJ
Sample : H4688-03 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SS-AST3

Quant Time: Sep 09 06:36:46 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF083116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 08 16:07:52 2016
Response via : Initial Calibration

