

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090817\
 Data File : BF098349.D
 Acq On : 8 Sep 2017 14:38
 Operator : SJ/JU
 Sample : I5110-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-1

Manual Integrations
 APPROVED

Sohil
 9/11/2017 4:47:31 PM

Quant Time: Sep 08 15:16:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 08 15:02:28 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	96676	20.00	ng	0.00
21) Naphthalene-d8	7.96	136	384548	20.00	ng	0.00
38) Acenaphthene-d10	9.72	164	178977	20.00	ng	0.00
63) Phenanthrene-d10	11.20	188	338846	20.00	ng	0.00
75) Chrysene-d12	13.83	240	209718	20.00	ng	0.00
86) Perylene-d12	15.24	264	189296	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	764598	120.30	ng	0.02
7) Phenol-d6	6.33	99	1045541	121.07	ng	0.00
23) Nitrobenzene-d5	7.25	82	619712	87.55	ng	0.00
41) 2,4,6-Tribromophenol	10.51	330	226394	145.79	ng	0.00
44) 2-Fluorobiphenyl	9.05	172	960146	78.82	ng	0.00
78) Terphenyl-d14	12.79	244	721179	71.71	ng	0.00
Target Compounds						
10) Phenol	6.35	94	36952	3.659	ng	88
49) Dimethylphthalate	9.44	163	117535	8.984	ng	98
70) Phenanthrene	11.22	178	39132	2.167	ng	99
74) Fluoranthene	12.40	202	66328	3.519	ng	98
77) Pyrene	12.63	202	56629	3.301	ng	97
87) Benzo(b)fluoranthene	14.84	252	28119m	2.357	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090817\
Data File : BF098349.D
Acq On : 8 Sep 2017 14:38
Operator : SJ/JU
Sample : I5110-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-1

Manual Integrations
APPROVED

Sohil
9/11/2017 4:47:31 PM

Quant Time: Sep 08 15:16:36 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 08 15:02:28 2017
Response via : Initial Calibration

