

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF090817\
 Data File : BF098365.D
 Acq On : 8 Sep 2017 22:03
 Operator : SJ/JU
 Sample : I5113-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RL-1-12.25-24.5

Manual Integrations
 APPROVED

Sohil
 9/11/2017 4:47:52 PM

Quant Time: Sep 09 02:53:19 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 17:28:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	96894	20.00	ng	-0.02
21) Naphthalene-d8	7.97	136	337894	20.00	ng	-0.02
38) Acenaphthene-d10	9.72	164	122426	20.00	ng	-0.02
63) Phenanthrene-d10	11.20	188	202965	20.00	ng	-0.02
75) Chrysene-d12	13.84	240	214309	20.00	ng	-0.02
86) Perylene-d12	15.25	264	154587	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	647896	101.71	ng	0.00
7) Phenol-d6	6.34	99	864658	99.90	ng	-0.01
23) Nitrobenzene-d5	7.25	82	508066	81.68	ng	-0.02
41) 2,4,6-Tribromophenol	10.51	330	111657	105.11	ng	-0.02
44) 2-Fluorobiphenyl	9.05	172	702647	84.32	ng	-0.02
78) Terphenyl-d14	12.79	244	503147	48.96	ng	-0.02
Target Compounds						
10) Phenol	6.35	94	36404	3.596	ng	96
31) Naphthalene	7.99	128	49811	2.840	ng	100
49) Dimethylphthalate	9.44	163	80222	8.964	ng	98
70) Phenanthrene	11.22	178	65738	6.078	ng	98
74) Fluoranthene	12.41	202	101366	8.979	ng	97
77) Pyrene	12.64	202	94670	5.401	ng	96
80) Benzo(a)anthracene	13.83	228	43580	3.393	ng	92
82) Chrysene	13.86	228	48230	3.775	ng	97
83) Bis(2-ethylhexyl)phthalate	13.82	149	20623	2.769	ng	# 99
87) Benzo(b)fluoranthene	14.85	252	46642m	4.787	ng	
89) Benzo(a)pyrene	15.19	252	30334	3.516	ng	# 72
91) Benzo(a,h,i)perylene	17.01	276	16998	2.320	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF090817\
Data File : BF098365.D
Acq On : 8 Sep 2017 22:03
Operator : SJ/JU
Sample : I5113-03
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RL-1-12.25-24.5

Manual Integrations
APPROVED

Sohil
9/11/2017 4:47:52 PM

Quant Time: Sep 09 02:53:19 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF081517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 17:28:20 2017
Response via : Initial Calibration

