

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090821\
 Data File : BF125398.D
 Acq On : 8 Sep 2021 13:11
 Operator : JU/CG
 Sample : PB138943BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB138943BL

Quant Time: Sep 08 14:15:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 08 14:14:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	171069	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	678516	20.000	ng	# 0.00
39) Acenaphthene-d10	9.927	164	368043	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	695981	20.000	ng	0.00
76) Chrysene-d12	14.057	240	477727	20.000	ng	# 0.00
86) Perylene-d12	15.545	264	381322	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1334810	118.102	ng	0.01
7) Phenol-d6	6.522	99	1672623	114.428	ng	0.00
23) Nitrobenzene-d5	7.451	82	1131974	83.989	ng	0.00
42) 2,4,6-Tribromophenol	10.721	330	558051	151.897	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	2016120	76.348	ng	0.00
79) Terphenyl-d14	13.004	244	2496245	83.243	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090821\
Data File : BF125398.D
Acq On : 8 Sep 2021 13:11
Operator : JU/CG
Sample : PB138943BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB138943BL

Quant Time: Sep 08 14:15:15 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 08 14:14:01 2021
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.886 min Scan# 816
Delta R.T. 0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11




Instrument :
BNA_F
ClientSampleId :
PB138943BL

Tgt Ion:152 Resp: 171069
Ion Ratio Lower Upper
152 100
150 158.5 117.6 176.4
115 59.3 47.1 70.7



#5
2-Fluorophenol
Concen: 118.102 ng
RT: 5.522 min Scan# 584
Delta R.T. 0.012 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion:112 Resp: 1334810
Ion Ratio Lower Upper
112 100
64 70.4 44.2 66.2#
63 38.3 26.2 39.2









#7
Phenol-d6
Concen: 114.428 ng
RT: 6.522 min Scan# 754
Delta R.T. -0.006 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11




Instrument :
BNA_F
ClientSampleId :
PB138943BL


Tgt Ion: 99 Resp: 1672623
Ion Ratio Lower Upper
99 100
42 20.9 23.5 35.3#
71 38.4 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion: 136 Resp: 678516
Ion Ratio Lower Upper
136 100
137 11.0 9.8 14.6
54 10.0 6.4 9.6#
68 6.0 3.3 4.9#









#23
Nitrobenzene-d5
Concen: 83.989 ng
RT: 7.451 min Scan# 912
Delta R.T. -0.006 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11




Instrument :
BNA_F
ClientSampleId :
PB138943BL

Tgt Ion: 82 Resp: 1131974
Ion Ratio Lower Upper
82 100
128 39.4 40.6 60.8#
54 60.8 49.6 74.4



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.927 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion: 164 Resp: 368043
Ion Ratio Lower Upper
164 100
162 104.7 81.3 121.9
160 45.8 38.0 57.0









#42
2,4,6-Tribromophenol
Concen: 151.897 ng
RT: 10.721 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11




Instrument :
BNA_F
ClientSampleId :
PB138943BL


Tgt Ion:330 Resp: 558051
Ion Ratio Lower Upper
330 100
332 95.9 76.2 114.4
141 27.6 44.6 66.8#



#45
2-Fluorobiphenyl
Concen: 76.348 ng
RT: 9.245 min Scan# 1217
Delta R.T. -0.006 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion:172 Resp: 2016120
Ion Ratio Lower Upper
172 100
171 35.8 26.5 39.7
170 24.3 18.8 28.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1586
Delta R.T. -0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11




Instrument :
BNA_F
ClientSampleId :
PB138943BL


Tgt Ion:188 Resp: 695981
Ion Ratio Lower Upper
188 100
94 8.2 6.7 10.1
80 9.2 7.5 11.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2035
Delta R.T. -0.006 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion:240 Resp: 477727
Ion Ratio Lower Upper
240 100
120 9.1 10.8 16.2#
236 25.7 21.6 32.4









#79
Terphenyl-d14
Concen: 83.243 ng
RT: 13.004 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Instrument :
BNA_F
ClientSampleId :
PB138943BL

Tgt Ion:244 Resp: 2496245
Ion Ratio Lower Upper
244 100
212 6.7 5.9 8.9
122 8.6 9.0 13.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.545 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BF125398.D
Acq: 8 Sep 2021 13:11

Tgt Ion:264 Resp: 381322
Ion Ratio Lower Upper
264 100
260 24.3 19.6 29.4
265 21.6 17.2 25.8

