

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF090822\  
 Data File : BF130177.D  
 Acq On : 08 Sep 2022 14:22  
 Operator : CG\JU  
 Sample : N4500-16  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SB-04-COMP

Quant Time: Sep 09 01:03:22 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF083122.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Sep 07 02:05:07 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.698  | 152  | 186304   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.975  | 136  | 721726   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.722  | 164  | 377437   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.204 | 188  | 585641   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 13.833 | 240  | 307231   | 20.000  | ng    | -0.01    |
| 86) Perylene-d12            | 15.233 | 264  | 361736   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.298  | 112  | 1175165  | 107.084 | ng    | 0.00     |
| 7) Phenol-d6                | 6.328  | 99   | 1452101  | 106.135 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.257  | 82   | 880002   | 78.360  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 10.510 | 330  | 313446   | 91.231  | ng    | -0.01    |
| 45) 2-Fluorobiphenyl        | 9.051  | 172  | 1646091  | 72.307  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.792 | 244  | 1394869  | 78.731  | ng    | 0.00     |

Target Compounds Qvalue

-----  
 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF090822\  
Data File : BF130177.D  
Acq On : 08 Sep 2022 14:22  
Operator : CG\JU  
Sample : N4500-16  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SB-04-COMP

Quant Time: Sep 09 01:03:22 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF083122.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Sep 07 02:05:07 2022  
Response via : Initial Calibration

