

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090822\
 Data File : BF130186.D
 Acq On : 08 Sep 2022 19:21
 Operator : CG\JU
 Sample : N4534-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FL-B

Quant Time: Sep 09 01:08:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 07 02:05:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.698	152	124949	20.000	ng	0.00
21) Naphthalene-d8	7.975	136	445237	20.000	ng	0.00
39) Acenaphthene-d10	9.728	164	215543	20.000	ng	0.00
64) Phenanthrene-d10	11.210	188	384922	20.000	ng	0.00
76) Chrysene-d12	13.874	240	218387	20.000	ng	# 0.03
86) Perylene-d12	15.286	264	260235	20.000	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.298	112	355290	48.272	ng	0.00
7) Phenol-d6	6.346	99	662512	72.201	ng	0.01
23) Nitrobenzene-d5	7.257	82	472996	68.273	ng	-0.01
42) 2,4,6-Tribromophenol	10.516	330	35379	18.032	ng	0.00
45) 2-Fluorobiphenyl	9.057	172	872933	67.146	ng	0.00
79) Terphenyl-d14	12.810	244	887044	70.437	ng	0.01
Target Compounds						
35) Caprolactam	8.410	113	5804	2.976	ng	Qvalue 98
74) Di-n-butylphthalate	11.780	149	77997	3.418	ng	# 94
84) Bis(2-ethylhexyl)phtha...	13.863	149	646317	66.737	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090822\
Data File : BF130186.D
Acq On : 08 Sep 2022 19:21
Operator : CG\JU
Sample : N4534-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FL-B

Quant Time: Sep 09 01:08:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF083122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 07 02:05:07 2022
Response via : Initial Calibration

